

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB165896BS

Contract: PARS02

Lab Code: CHEM Case No.: P5362

SAS No.: P5362

SDG NO.: P5362

Lab Sample ID: PB165896BS

Date(s) Analyzed: 12/27/2024 12/27/2024

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column:(2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.33	8.28	8.38	5.30	7.8
	2	8.93	8.88	8.98	4.90	
2,4,5-TP(Silvex)	1	9.20	9.15	9.25	5.60	11.3
	2	9.83	9.78	9.88	5.00	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

WC-SOIL-20241219MS

Contract: PARS02

Lab Code: CHEM Case No.: P5362

SAS No.: P5362 SDG NO.: P5362

Lab Sample ID: P5362-02MS

Date(s) Analyzed: 12/27/2024 12/27/2024

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column:(2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.32	8.27	8.37	56.3	10.5
	2	8.93	8.88	8.98	50.7	
2,4,5-TP(Silvex)	1	9.20	9.15	9.25	56.3	7.4
	2	9.84	9.79	9.89	60.6	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

WC-SOIL-20241219MSD

Contract: PARS02

Lab Code: CHEM **Case No.:** P5362 **SAS No.:** P5362 **SDG NO.:** P5362

Lab Sample ID: P5362-02MSD **Date(s) Analyzed:** 12/27/2024 12/27/2024

Instrument ID (1): ECD_S **Instrument ID (2):** ECD_S

GC Column: (1): RTX-CLP **ID:** 0.32 (mm) **GC Column:(2):** RTX-CLP2 **ID:** 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.33	8.28	8.38	55.1	8.9
	2	8.93	8.88	8.98	50.4	
2,4,5-TP(Silvex)	1	9.20	9.15	9.25	56.3	10.4
	2	9.84	9.79	9.89	62.5	