

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: PARS02

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

Instrument ID: ECD_S Calibration Date(s): 12/23/2024 12/23/2024
Calibration Times: 11:23 12:59

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS028787.D</u>	CF 500 =	<u>PS028788.D</u>		
CF 750 =		<u>PS028789.D</u>	CF 1000 =	<u>PS028790.D</u>	CF 1500 =	<u>PS028791.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	16812700000	15996100000	15801400000	15298000000	14638400000	15709300000	5
2,4-D	3146520000	2896320000	2842430000	2725340000	2637060000	2849530000	7
2,4-DCAA	2519360000	2280310000	2254700000	2149750000	2106010000	2262030000	7

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: PARS02

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

Instrument ID: ECD_S Calibration Date(s): 12/23/2024 12/23/2024
Calibration Times: 11:23 12:59

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS028787.D</u>	CF 500 =	<u>PS028788.D</u>		
CF 750 =		<u>PS028789.D</u>	CF 1000 =	<u>PS028790.D</u>	CF 1500 =	<u>PS028791.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	9147430000	9053780000	9109830000	8920050000	8686470000	8983510000	2
2,4-D	1472990000	1449810000	1452790000	1429890000	1410250000	1443150000	2
2,4-DCAA	1184820000	1126220000	1123240000	1104090000	1089310000	1125540000	3