

### CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL01 Date Analyzed: 12/27/2024

Lab Sample No.: HSTDCCC750 Data File : PS028837.D Time Analyzed: 19:56

| COMPOUND         | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|-------|-----------|-------|--------------------|-------------------|------|
|                  |       | FROM      | TO    |                    |                   |      |
| 2,4,5-TP(Silvex) | 9.198 | 9.106     | 9.306 | 789.440            | 712.500           | 10.8 |
| 2,4-D            | 8.325 | 8.232     | 8.432 | 737.600            | 705.000           | 4.6  |
| 2,4-DCAA         | 7.205 | 7.110     | 7.310 | 805.920            | 750.000           | 7.5  |

### CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL01 Date Analyzed: 12/27/2024

Lab Sample No.: HSTDCCC750 Data File : PS028837.D Time Analyzed: 19:56

| COMPOUND         | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D  |
|------------------|-------|-----------|-------|--------------------|-------------------|-----|
|                  |       | FROM      | TO    |                    |                   |     |
| 2,4,5-TP(Silvex) | 9.834 | 9.742     | 9.942 | 729.080            | 712.500           | 2.3 |
| 2,4-D            | 8.932 | 8.840     | 9.040 | 705.540            | 705.000           | 0.1 |
| 2,4-DCAA         | 7.694 | 7.600     | 7.800 | 754.200            | 750.000           | 0.6 |