

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL02 Date Analyzed: 12/28/2024

Lab Sample No.: HSTDCCC750 Data File : PS028846.D Time Analyzed: 00:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.198	9.106	9.306	801.390	712.500	12.5
2,4-D	8.324	8.232	8.432	742.400	705.000	5.3
2,4-DCAA	7.205	7.110	7.310	815.240	750.000	8.7

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL02 Date Analyzed: 12/28/2024

Lab Sample No.: HSTDCCC750 Data File : PS028846.D Time Analyzed: 00:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.833	9.742	9.942	739.050	712.500	3.7
2,4-D	8.931	8.840	9.040	707.040	705.000	0.3
2,4-DCAA	7.693	7.600	7.800	763.530	750.000	1.8