

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL03 Date Analyzed: 12/31/2024

Lab Sample No.: HSTDCCC750 Data File : PS028862.D Time Analyzed: 08:59

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.201	9.106	9.306	765.240	712.500	7.4
2,4-D	8.327	8.232	8.432	742.010	705.000	5.2
2,4-DCAA	7.207	7.110	7.310	790.330	750.000	5.4

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL03 Date Analyzed: 12/31/2024

Lab Sample No.: HSTDCCC750 Data File : PS028862.D Time Analyzed: 08:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.828	9.742	9.942	678.000	712.500	-4.8
2,4-D	8.928	8.840	9.040	660.870	705.000	-6.3
2,4-DCAA	7.690	7.600	7.800	691.610	750.000	-7.8