

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM **Case No.:** P5362 **SAS No.:** P5362 **SDG NO.:** P5362

GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/23/2024 12/23/2024

Client Sample No.: CCAL04 **Date Analyzed:** 12/31/2024

Lab Sample No.: HSTDCCC750 **Data File :** PS028865.D **Time Analyzed:** 12:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.202	9.106	9.306	780.130	712.500	9.5
2,4-D	8.328	8.232	8.432	750.400	705.000	6.4
2,4-DCAA	7.208	7.110	7.310	803.900	750.000	7.2

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL04 Date Analyzed: 12/31/2024

Lab Sample No.: HSTDCCC750 Data File : PS028865.D Time Analyzed: 12:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.829	9.742	9.942	692.120	712.500	-2.9
2,4-D	8.928	8.840	9.040	674.020	705.000	-4.4
2,4-DCAA	7.691	7.600	7.800	708.530	750.000	-5.5