

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: PARS02

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

Instrument ID: ECD_L

Calibration Date(s): 12/23/2024 12/23/2024

Calibration Times: 13:15 14:09

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PL093484.D</u>	CF 075 =	<u>PL093485.D</u>		
CF 050 =		<u>PL093486.D</u>	CF 025 =	<u>PL093487.D</u>	CF 005 =	<u>PL093488.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
Decachlorobiphenyl	1661260000	1649170000	1775440000	1867730000	2291490000	1849020000	14
Endrin	1980610000	1930990000	2046720000	2145760000	2658130000	2152440000	14
gamma-BHC (Lindane)	3189620000	3052290000	3180150000	3224710000	3750380000	3279430000	8
Heptachlor	2746960000	2670120000	2802230000	2919950000	3502290000	2928310000	11
Heptachlor epoxide	2426470000	2365400000	2505620000	2624560000	3249710000	2634350000	14
Methoxychlor	902299000	897910000	965987000	1022970000	1209220000	999678000	13
Tetrachloro-m-xylene	2318290000	2256280000	2391520000	2493110000	2919250000	2475690000	11

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: PARS02

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

Instrument ID: ECD_L Calibration Date(s): 12/23/2024 12/23/2024

Calibration Times: 13:15 14:09

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PL093484.D</u>	CF 075 =	<u>PL093485.D</u>		
CF 050 =		<u>PL093486.D</u>	CF 025 =	<u>PL093487.D</u>	CF 005 =	<u>PL093488.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
Decachlorobiphenyl	2956580000	2818470000	2885080000	2902790000	3366620000	2985910000	7
Endrin	3457020000	3288270000	3315120000	3193900000	3287350000	3308330000	3
gamma-BHC (Lindane)	4493780000	4228750000	4260470000	4062450000	4051390000	4219370000	4
Heptachlor	4294490000	4097760000	4159520000	4029940000	4197880000	4155920000	2
Heptachlor epoxide	3890040000	3699480000	3757900000	3697650000	4098450000	3828700000	4
Methoxychlor	1579690000	1538710000	1589880000	1595910000	1744000000	1609640000	5
Tetrachloro-m-xylene	2947220000	2813690000	2902100000	2859340000	3033640000	2911200000	3