

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL01 Date Analyzed: 12/27/2024

Lab Sample No.: PSTDCCC050 Data File : PL093541.D Time Analyzed: 14:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.054	8.956	9.156	51.650	50.000	3.3
Endrin	6.573	6.475	6.675	50.410	50.000	0.8
gamma-BHC (Lindane)	4.327	4.229	4.429	53.260	50.000	6.5
Heptachlor	4.916	4.818	5.018	52.330	50.000	4.7
Heptachlor epoxide	5.684	5.586	5.786	52.340	50.000	4.7
Methoxychlor	7.500	7.400	7.600	53.550	50.000	7.1
Tetrachloro-m-xylene	3.539	3.442	3.642	52.590	50.000	5.2

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL01 Date Analyzed: 12/27/2024

Lab Sample No.: PSTDCCC050 Data File : PL093541.D Time Analyzed: 14:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.912	7.812	8.012	53.120	50.000	6.2
Endrin	5.638	5.540	5.740	54.880	50.000	9.8
gamma-BHC (Lindane)	3.608	3.510	3.710	55.410	50.000	10.8
Heptachlor	3.946	3.848	4.048	54.690	50.000	9.4
Heptachlor epoxide	4.729	4.630	4.830	54.170	50.000	8.3
Methoxychlor	6.611	6.512	6.712	54.610	50.000	9.2
Tetrachloro-m-xylene	2.774	2.677	2.877	54.120	50.000	8.2