

SOP ID: M3510C,3580A-Extraction Pesticide-16

Clean Up SOP #: N/A

Matrix : Water

Weigh By: RJ

Balance check: N/A

Balance ID: N/A

pH Strip Lot#: N/A

Extraction By: RJ

Filter By: RJ

pH Meter ID: N/A

Hood ID: 4,6,7

Extraction Start Date : 12/27/2024

Extraction Start Time : 11:00

Extraction End Date : 12/27/2024

Extraction End Time : 15:55

Concentration By: EH

Supervisor By : rajesh

Extraction Method: ☒ Seperatory Funnel ☐ Continious Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet

Standarded Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	500 PPB	PP23928
Surrogate	1.0ML	200 PPB	PP23985
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane	N/A	E3847
Baked Na2SO4	N/A	EP2573
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40 ML Vial lot# 03-40 BTS721.

KD Bath ID: WATER BATH-1,2

KD Bath Temperature: 60 °C

Envap ID: NEVAP-02

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
12/27/24	Rp (Ext. Lab)	Rp (Ext. Lab)
16:00	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction Pesticide-16

Concentration Date: 12/27/2024

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB165858TB	PB165858TB	TCLP Pesticide	100	6	ritesh	rajesh	10			SEP-01
PB165895BL	PBLK895	TCLP Pesticide	1000	6	ritesh	rajesh	10			2
PB165895BS	PLCS895	TCLP Pesticide	1000	6	ritesh	rajesh	10			3
P5362-02	WC-SOIL-20241219	TCLP Pesticide	100	6	ritesh	rajesh	10	A		4
P5362-02MS	WC-SOIL-20241219MS	TCLP Pesticide	100	6	ritesh	rajesh	10	A		5
P5362-02MS D	WC-SOIL-20241219MSD	TCLP Pesticide	100	6	ritesh	rajesh	10	A		6
P5380-02	TAPIAL3-IDW-SOIL-1220 24-T1	TCLP Pesticide	100	6	ritesh	rajesh	10	A		7

* Extracts relinquished on the same date as received.

[Signature]
12/27/24