

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: PARS02
 ProjectID: Con Edison Non-MGP - East River 453648.60024.03
 Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG No.: P5362

Calibration Sequence : FG121924		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	200331626	117842	FG014987.D
850	106454179	125240	FG014988.D
340	43934817	129220	FG014989.D
170	26196418	154097	FG014990.D
85	12067151	141966	FG014991.D
AVG RF : 133673		% RSD : 10.76	AVG RT : 15.074