

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG No.: P5362
Instrument ID: BNA_F Calibration Date(s): 12/26/2024 12/26/2024
Calibration Time(s): 16:23 19:25

LAB FILE ID:		RRF2.5 = BF140970.D			RRF005 = BF140971.D		RRF010 = BF140972.D	
		RRF020 = BF140973.D			RRF040 = BF140974.D		RRF050 = BF140975.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.502	1.422	1.423	1.293	1.258	1.327	9.4
Benzaldehyde			1.193	1.089	0.861	0.792	0.941	20.2
Phenol-d6		1.910	1.839	1.811	1.667	1.625	1.709	8.6
Phenol		1.956	1.880	1.866	1.744	1.683	1.767	8.0
bis(2-Chloroethyl) ether		1.543	1.479	1.449	1.363	1.416	1.427	4.8
2-Chlorophenol		1.556	1.484	1.483	1.386	1.337	1.399	8.2
2-Methylphenol		1.250	1.208	1.226	1.151	1.125	1.162	5.8
2,2-oxybis(1-Chloropropane)		2.420	2.317	2.313	2.112	2.048	2.158	9.0
Acetophenone		0.575	0.535	0.521	0.470	0.459	0.491	11.0
3+4-Methylphenols		1.644	1.624	1.598	1.452	1.387	1.471	10.6
n-Nitroso-di-n-propylamine	1.133	1.140	1.127	1.079	1.008	0.988	1.047	8.0
Nitrobenzene-d5		0.244	0.277	0.321	0.327	0.338	0.313	12.2
Hexachloroethane		0.587	0.573	0.584	0.544	0.538	0.553	5.3
Nitrobenzene		0.288	0.319	0.364	0.367	0.371	0.349	9.5
Isophorone		0.744	0.706	0.714	0.657	0.654	0.680	6.1
2-Nitrophenol		0.067	0.080	0.107	0.127	0.142	0.118	29.3
2,4-Dimethylphenol		0.268	0.255	0.260	0.241	0.241	0.249	5.2
bis(2-Chloroethoxy) methane		0.477	0.462	0.462	0.415	0.412	0.431	8.3
2,4-Dichlorophenol		0.286	0.289	0.295	0.278	0.281	0.282	3.4
Naphthalene		1.209	1.154	1.142	1.020	0.989	1.054	11.0
4-Chloroaniline		0.414	0.398	0.404	0.367	0.368	0.381	6.2
Hexachlorobutadiene		0.208	0.199	0.196	0.182	0.181	0.188	7.1
Caprolactam		0.098	0.095	0.100	0.094	0.094	0.096	2.4
4-Chloro-3-methylphenol		0.330	0.320	0.327	0.309	0.304	0.313	4.2
2-Methylnaphthalene		0.772	0.740	0.732	0.642	0.629	0.674	10.8
Hexachlorocyclopentadiene		0.165	0.176	0.189	0.187	0.189	0.182	5.0
2,4,6-Trichlorophenol		0.354	0.356	0.398	0.364	0.356	0.366	4.6
2-Fluorobiphenyl		1.546	1.452	1.378	1.175	1.129	1.255	16.1
2,4,5-Trichlorophenol		0.371	0.389	0.377	0.380	0.386	0.375	3.2
1,1-Biphenyl		1.776	1.696	1.651	1.479	1.421	1.531	11.6
2-Chloronaphthalene		1.362	1.293	1.270	1.140	1.115	1.188	10.2
2-Nitroaniline		0.139	0.182	0.248	0.288	0.306	0.259	28.5
Dimethylphthalate		1.438	1.414	1.389	1.285	1.246	1.316	7.4
Acenaphthylene		1.934	1.869	1.824	1.654	1.619	1.710	9.8

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG No.: P5362
Instrument ID: BNA_F Calibration Date(s): 12/26/2024 12/26/2024
Calibration Time(s): 16:23 19:25

LAB FILE ID:		RRF2.5 = BF140970.D			RRF005 = BF140971.D		RRF010 = BF140972.D	
		RRF020 = BF140973.D			RRF040 = BF140974.D		RRF050 = BF140975.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.128	0.173	0.233	0.256	0.269	0.231	25.2
3-Nitroaniline		0.153	0.197	0.260	0.281	0.289	0.255	22.6
Acenaphthene		1.289	1.240	1.229	1.119	1.096	1.156	8.4
2,4-Dinitrophenol			0.037	0.053	0.072	0.086	0.075	35.0
4-Nitrophenol			0.166	0.213	0.229	0.237	0.225	14.4
Dibenzofuran		1.857	1.780	1.743	1.565	1.523	1.624	10.5
2,4-Dinitrotoluene		0.136	0.187	0.265	0.310	0.326	0.273	30.1
Diethylphthalate		1.433	1.386	1.369	1.266	1.230	1.299	7.5
4-Chlorophenyl-phenylether		0.700	0.673	0.653	0.579	0.560	0.606	11.3
Fluorene		1.497	1.432	1.351	1.192	1.146	1.260	13.2
4-Nitroaniline		0.156	0.203	0.257	0.278	0.290	0.255	21.7
4,6-Dinitro-2-methylphenol			0.033	0.051	0.066	0.078	0.067	32.7
n-Nitrosodiphenylamine		0.712	0.675	0.668	0.605	0.594	0.630	8.8
2,4,6-Tribromophenol		0.177	0.189	0.203	0.196	0.198	0.194	4.6
4-Bromophenyl-phenylether		0.235	0.229	0.228	0.213	0.212	0.216	7.0
Hexachlorobenzene		0.261	0.247	0.249	0.232	0.231	0.238	6.0
Atrazine		0.218	0.205	0.189	0.156	0.147	0.176	18.5
Pentachlorophenol		0.122	0.137	0.154	0.155	0.156	0.148	9.4
Phenanthrene		1.221	1.168	1.129	1.005	0.982	1.053	11.3
Anthracene		1.187	1.136	1.113	0.991	0.963	1.031	11.0
Carbazole		1.138	1.093	1.076	0.947	0.923	0.992	11.0
Di-n-butylphthalate		1.248	1.238	1.237	1.083	1.054	1.128	9.7
Fluoranthene		1.333	1.274	1.242	1.091	1.050	1.142	12.3
Pyrene		1.754	1.670	1.701	1.692	1.783	1.749	3.6
Terphenyl-d14		1.275	1.181	1.188	1.141	1.194	1.204	3.6
Butylbenzylphthalate		0.532	0.571	0.640	0.656	0.684	0.639	10.1
3,3-Dichlorobenzidine		0.474	0.455	0.455	0.403	0.382	0.419	9.8
Benzo (a) anthracene		1.573	1.452	1.437	1.369	1.340	1.394	7.4
Chrysene		1.341	1.330	1.327	1.208	1.194	1.257	5.7
Bis (2-ethylhexyl) phthalate		0.822	0.842	0.859	0.820	0.827	0.828	2.5
Di-n-octyl phthalate		1.157	1.231	1.253	1.199	1.170	1.199	2.8
Benzo (b) fluoranthene		1.360	1.564	1.479	1.417	1.306	1.395	7.2
Benzo (k) fluoranthene		1.165	1.121	1.301	1.065	1.073	1.093	11.2
Benzo (a) pyrene		1.146	1.108	1.137	1.069	1.047	1.083	4.6
Indeno (1,2,3-cd) pyrene		1.050	1.009	1.224	1.311	1.397	1.260	13.7

All other compounds must meet a minimum RRF of 0.010.

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG No.: P5362
Instrument ID: BNA_F Calibration Date(s): 12/26/2024 12/26/2024
Calibration Time(s): 16:23 19:25

LAB FILE ID:		RRF2.5 = BF140970.D		RRF005 = BF140971.D		RRF010 = BF140972.D		
		RRF020 = BF140973.D		RRF040 = BF140974.D		RRF050 = BF140975.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo (a,h) anthracene		0.847	0.829	0.992	1.070	1.127	1.021	13.4
Benzo (g,h,i) perylene		0.825	0.811	1.023	1.126	1.194	1.059	16.9
1,2,4,5-Tetrachlorobenzene		0.602	0.586	0.586	0.535	0.527	0.550	7.5
1,4-Dioxane		0.630	0.621	0.624	0.597	0.587	0.601	4.3
2,3,4,6-Tetrachlorophenol		0.303	0.303	0.327	0.314	0.307	0.310	3.0