

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122724\
 Data File : BF140994.D
 Acq On : 27 Dec 2024 13:59
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 27 14:47:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 27 00:31:16 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	109	0.00
2	1,4-Dioxane	40.000	39.280	1.8	108	0.00
3	Pyridine	40.000	40.285	-0.7	112	0.00
4	n-Nitrosodimethylamine	40.000	39.553	1.1	108	0.00
5 S	2-Fluorophenol	80.000	76.459	4.4	107	0.00
6	Aniline	40.000	40.582	-1.5	104	-0.02
7 S	Phenol-d6	80.000	76.023	5.0	106	0.00
8	2-Chlorophenol	40.000	38.720	3.2	107	0.00
9	Benzaldehyde	40.000	44.682	-11.7	116	0.00
10 C	Phenol	40.000	38.488	3.8	106	0.00
11	bis(2-Chloroethyl)ether	40.000	36.923	7.7	105	0.00
12	1,3-Dichlorobenzene	40.000	38.974	2.6	108	0.00
13 C	1,4-Dichlorobenzene	40.000	38.694	3.3	108	0.00
14	1,2-Dichlorobenzene	40.000	38.312	4.2	107	0.00
15	Benzyl Alcohol	40.000	37.984	5.0	104	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.887	5.3	106	0.00
17	2-Methylphenol	40.000	38.327	4.2	106	0.00
18	Hexachloroethane	40.000	39.154	2.1	109	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.927	7.7	105	0.00
20	3+4-Methylphenols	40.000	38.461	3.8	106	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	106	0.00
22	Acetophenone	40.000	38.002	5.0	106	0.00
23 S	Nitrobenzene-d5	80.000	91.990	-15.0	117	0.00
24	Nitrobenzene	40.000	44.733	-11.8	113	0.00
25	Isophorone	40.000	38.034	4.9	105	0.00
26 C	2-Nitrophenol	40.000	45.034	-12.6	134	0.00
27	2,4-Dimethylphenol	40.000	39.199	2.0	107	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.937	5.2	105	0.00
29 C	2,4-Dichlorophenol	40.000	39.596	1.0	107	0.00
30	1,2,4-Trichlorobenzene	40.000	39.171	2.1	106	0.00
31	Naphthalene	40.000	38.456	3.9	106	0.00
32	Benzoic acid	40.000	43.059	-7.6	129	0.00
33	4-Chloroaniline	40.000	37.952	5.1	105	0.00
34 C	Hexachlorobutadiene	40.000	38.912	2.7	107	0.00
35	Caprolactam	40.000	39.233	1.9	106	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.663	3.3	104	0.00
37	2-Methylnaphthalene	40.000	38.321	4.2	107	0.00
38	1-Methylnaphthalene	40.000	37.754	5.6	105	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.245	1.9	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.751	-1.9	105	0.00
42 S	2,4,6-Tribromophenol	80.000	84.432	-5.5	110	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.100	-0.3	106	0.00
44	2,4,5-Trichlorophenol	40.000	41.204	-3.0	107	0.00
45 S	2-Fluorobiphenyl	80.000	75.089	6.1	106	0.00
46	1,1'-Biphenyl	40.000	38.706	3.2	106	0.00
47	2-Chloronaphthalene	40.000	38.378	4.1	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122724\
 Data File : BF140994.D
 Acq On : 27 Dec 2024 13:59
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 27 14:47:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 27 00:31:16 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.257	-13.1	128	0.00
49	Acenaphthylene	40.000	38.857	2.9	106	0.00
50	Dimethylphthalate	40.000	38.961	2.6	105	0.00
51	2,6-Dinitrotoluene	40.000	44.477	-11.2	122	0.00
52 C	Acenaphthene	40.000	38.833	2.9	106	0.00
53	3-Nitroaniline	40.000	44.203	-10.5	120	0.00
54 P	2,4-Dinitrophenol	40.000	47.390	-18.5	146	0.00
55	Dibenzofuran	40.000	38.499	3.8	105	0.00
56 P	4-Nitrophenol	40.000	43.761	-9.4	113	0.00
57	2,4-Dinitrotoluene	40.000	46.211	-15.5	129	0.00
58	Fluorene	40.000	37.810	5.5	105	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.652	-1.6	106	0.00
60	Diethylphthalate	40.000	38.843	2.9	105	0.00
61	4-Chlorophenyl-phenylether	40.000	37.721	5.7	104	0.00
62	4-Nitroaniline	40.000	43.704	-9.3	120	0.00
63	Azobenzene	40.000	38.398	4.0	105	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.917	-17.3	143	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.161	2.1	104	0.00
67	4-Bromophenyl-phenylether	40.000	40.629	-1.6	105	0.00
68	Hexachlorobenzene	40.000	39.462	1.3	104	0.00
69	Atrazine	40.000	35.956	10.1	103	0.00
70 C	Pentachlorophenol	40.000	43.264	-8.2	106	0.00
71	Phenanthrene	40.000	39.276	1.8	105	0.00
72	Anthracene	40.000	38.909	2.7	103	0.00
73	Carbazole	40.000	38.935	2.7	104	0.00
74	Di-n-butylphthalate	40.000	39.149	2.1	104	0.00
75 C	Fluoranthene	40.000	37.866	5.3	101	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzydine	40.000	50.365	-25.9#	129	0.00
78	Pyrene	40.000	38.653	3.4	100	0.00
79 S	Terphenyl-d14	80.000	77.455	3.2	102	0.00
80	Butylbenzylphthalate	40.000	40.421	-1.1	99	0.00
81	Benzo(a)anthracene	40.000	37.774	5.6	96	0.00
82	3,3'-Dichlorobenzidine	40.000	38.791	3.0	101	0.00
83	Chrysene	40.000	39.340	1.6	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.580	1.1	100	0.00
85 c	Di-n-octyl phthalate	40.000	40.727	-1.8	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	108	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	40.992	-2.5	107	-0.01
88	Benzo(b)fluoranthene	40.000	35.555	11.1	95	0.00
89	Benzo(k)fluoranthene	40.000	44.855	-12.1	125	0.00
90 C	Benzo(a)pyrene	40.000	39.424	1.4	108	-0.01
91	Dibenzo(a,h)anthracene	40.000	40.991	-2.5	106	0.00
92	Benzo(g,h,i)perylene	40.000	39.931	0.2	102	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122724\
Data File : BF140994.D
Acq On : 27 Dec 2024 13:59
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 27 14:47:17 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 27 00:31:16 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0