

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5362

Client: PARSONS Main of New York, Inc.

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: P5362-01MS		Client Sample ID: WC-SOIL-20241219MS		DataFile: BF141002.D							
Benzaldehyde	1800	0	490	ug/Kg	27				10	86	
Phenol	1800	0	1900	ug/Kg	106				67	126	
bis(2-Chloroethyl)ether	1800	0	1700	ug/Kg	94				54	125	
2-Chlorophenol	1800	0	1900	ug/Kg	106				79	107	
2-Methylphenol	1800	0	1900	ug/Kg	106				66	122	
2,2-oxybis(1-Chloropropane)	1800	0	1800	ug/Kg	100				65	110	
Acetophenone	1800	0	1700	ug/Kg	94				75	111	
3+4-Methylphenols	1800	0	1800	ug/Kg	100				66	104	
N-Nitroso-di-n-propylamine	1800	0	1800	ug/Kg	100				59	119	
Hexachloroethane	1800	0	1800	ug/Kg	100				65	117	
Nitrobenzene	1800	0	2000	ug/Kg	111				70	119	
Isophorone	1800	0	1800	ug/Kg	100				76	122	
2-Nitrophenol	1800	0	2200	ug/Kg	122				54	145	
2,4-Dimethylphenol	1800	0	2200	ug/Kg	122				44	135	
bis(2-Chloroethoxy)methane	1800	0	1800	ug/Kg	100				68	112	
2,4-Dichlorophenol	1800	0	1900	ug/Kg	106				72	118	
Naphthalene	1800	0	1700	ug/Kg	94				72	110	
4-Chloroaniline	1800	0	350	ug/Kg	19				10	91	
Hexachlorobutadiene	1800	0	1700	ug/Kg	94				66	114	
Caprolactam	1800	0	1800	ug/Kg	100				51	134	
4-Chloro-3-methylphenol	1800	0	1900	ug/Kg	106				57	132	
2-Methylnaphthalene	1800	0	1800	ug/Kg	100				59	123	
Hexachlorocyclopentadiene	3600	0	4800	ug/Kg	133				10	175	
2,4,6-Trichlorophenol	1800	0	1900	ug/Kg	106				72	117	
2,4,5-Trichlorophenol	1800	0	1900	ug/Kg	106				72	117	
1,1-Biphenyl	1800	0	1800	ug/Kg	100				75	113	
2-Chloronaphthalene	1800	0	1700	ug/Kg	94				67	118	
2-Nitroaniline	1800	0	2100	ug/Kg	117				69	127	
Dimethylphthalate	1800	0	1900	ug/Kg	106				70	113	
Acenaphthylene	1800	0	1900	ug/Kg	106				79	118	
2,6-Dinitrotoluene	1800	0	2000	ug/Kg	111				70	125	
3-Nitroaniline	1800	0	990	ug/Kg	55				30	99	
Acenaphthene	1800	0	2000	ug/Kg	111				70	121	
2,4-Dinitrophenol	3600	0	3500	ug/Kg	97				10	155	
4-Nitrophenol	3600	0	3800	ug/Kg	106				45	133	
Dibenzofuran	1800	0	1800	ug/Kg	100				72	110	
2,4-Dinitrotoluene	1800	0	2100	ug/Kg	117				55	128	
Diethylphthalate	1800	0	1800	ug/Kg	100				70	112	
4-Chlorophenyl-phenylether	1800	0	1800	ug/Kg	100				71	108	
Fluorene	1800	0	1700	ug/Kg	94				68	116	
4-Nitroaniline	1800	0	1700	ug/Kg	94				55	120	
4,6-Dinitro-2-methylphenol	1800	0	2200	ug/Kg	122				10	160	
N-Nitrosodiphenylamine	1800	0	1900	ug/Kg	106				73	118	
4-Bromophenyl-phenylether	1800	0	2000	ug/Kg	111				65	121	
Hexachlorobenzene	1800	0	1900	ug/Kg	106				67	118	
Atrazine	1800	0	2300	ug/Kg	128	*			79	127	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5362

Client: PARSONS Main of New York, Inc.

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Pentachlorophenol	3600	0	3700	ug/Kg	103				47	128	
Phenanthrene	1800	0	1900	ug/Kg	106				52	128	
Anthracene	1800	0	1900	ug/Kg	106				62	124	
Carbazole	1800	0	1700	ug/Kg	94				59	119	
Di-n-butylphthalate	1800	0	1900	ug/Kg	106				69	118	
Fluoranthene	1800	0	1600	ug/Kg	89				44	125	
Pyrene	1800	92.7	1900	ug/Kg	100				26	142	
Butylbenzylphthalate	1800	0	1900	ug/Kg	106				64	126	
3,3-Dichlorobenzidine	1800	0	1200	ug/Kg	67				33	116	
Benzo(a)anthracene	1800	0	1700	ug/Kg	94				71	114	
Chrysene	1800	0	1900	ug/Kg	106				57	121	
bis(2-Ethylhexyl)phthalate	1800	0	1900	ug/Kg	106				42	169	
Di-n-octyl phthalate	1800	0	2200	ug/Kg	122				23	175	
Benzo(b)fluoranthene	1800	97.9	1700	ug/Kg	89				67	121	
Benzo(k)fluoranthene	1800	0	2100	ug/Kg	117				57	134	
Benzo(a)pyrene	1800	150	2100	ug/Kg	108				70	142	
Indeno(1,2,3-cd)pyrene	1800	0	1800	ug/Kg	100				40	129	
Dibenz(a,h)anthracene	1800	0	1800	ug/Kg	100				43	123	
Benzo(g,h,i)perylene	1800	150	1600	ug/Kg	81				24	125	
1,2,4,5-Tetrachlorobenzene	1800	0	1800	ug/Kg	100				69	124	
1,4-Dioxane	1800	0	1700	ug/Kg	94				46	112	
2,3,4,6-Tetrachlorophenol	1800	0	2000	ug/Kg	111				69	112	