

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

EPA Sample No.: SSTDCCC040 Date Analyzed: 12/27/2024

Lab File ID: BF140981.D Time Analyzed: 08:19

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	281239	6.828	1080100	8.11	567563	9.86
UPPER LIMIT	562478	7.328	2160200	8.61	1135130	10.363
LOWER LIMIT	140620	6.328	540050	7.61	283782	9.363
EPA SAMPLE NO.						
01 PB165814BL	273904	6.83	1090160	8.10	581774	9.86
02 PB165814BS	251701	6.83	1024350	8.11	541141	9.86

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.