

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

EPA Sample No.: SSTDCCC040 Date Analyzed: 12/27/2024

Lab File ID: BF140981.D Time Analyzed: 08:19

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1005110	11.345	657302	13.986	544134	15.439
UPPER LIMIT	2010220	11.845	1314600	14.486	1088270	15.939
LOWER LIMIT	502555	10.845	328651	13.486	272067	14.939
EPA SAMPLE NO.						
01 PB165814BL	982068	11.35	752306	13.98	594165	15.44
02 PB165814BS	919053	11.35	633131	13.99	527325	15.44

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.