

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PARS02

Lab Code: CHEM

SAS No.: P5362

SDG NO.: P5362

Lab File ID: BF140980.D

DFTPP Injection Date: 12/27/2024

Instrument ID: BNA_F

DFTPP Injection Time: 07:54

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	37.9
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	35.9
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	47.8
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	28.1
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	14.8
442	Greater than 50% of mass 198	94.4
443	15.0 - 24.0% of mass 442	18.3 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140981.D	12/27/2024	08:19
PB165814BL	PB165814BL	BF140989.D	12/27/2024	11:48
PB165814BS	PB165814BS	BF140990.D	12/27/2024	12:14