

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5362

Client: PARSONS Main of New York, Inc.

Analytical Method: 8270E DataFile: BF141021.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB165894BS	Pyridine	50	45.5	ug/L	91				29	97	
	1,4-Dichlorobenzene	50	47.7	ug/L	95				76	103	
	2-Methylphenol	50	50.1	ug/L	100				69	109	
	3+4-Methylphenols	50	48.9	ug/L	98				67	106	
	Hexachloroethane	50	49.2	ug/L	98				76	118	
	Nitrobenzene	50	51.7	ug/L	103				58	106	
	Hexachlorobutadiene	50	48.2	ug/L	96				69	101	
	2,4,6-Trichlorophenol	50	52.9	ug/L	106				61	110	
	2,4,5-Trichlorophenol	50	53.4	ug/L	107		*		70	106	
	2,4-Dinitrotoluene	50	55.4	ug/L	111				60	115	
	Hexachlorobenzene	50	54.3	ug/L	109		*		73	106	
	Pentachlorophenol	100	110	ug/L	110				47	114	