

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PARS02

Lab Code: CHEM

SAS No.: P5362 SDG NO.: P5362

Lab File ID: BF140969.D

DFTPP Injection Date: 12/26/2024

Instrument ID: BNA_F

DFTPP Injection Time: 15:57

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	38.4
68	Less than 2.0% of mass 69	0.6 (1.7) 1
69	Mass 69 relative abundance	35.5
70	Less than 2.0% of mass 69	0.2 (0.4) 1
127	10.0 - 80.0% of mass 198	47.8
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	27.2
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	14.3
442	Greater than 50% of mass 198	90.5
443	15.0 - 24.0% of mass 442	17.8 (19.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF140970.D	12/26/2024	16:23
SSTDICC005	SSTDICC005	BF140971.D	12/26/2024	16:49
SSTDICC010	SSTDICC010	BF140972.D	12/26/2024	17:15
SSTDICC020	SSTDICC020	BF140973.D	12/26/2024	17:41
SSTDICCC040	SSTDICCC040	BF140974.D	12/26/2024	18:06
SSTDICC050	SSTDICC050	BF140975.D	12/26/2024	18:33
SSTDICC060	SSTDICC060	BF140976.D	12/26/2024	18:59
SSTDICC080	SSTDICC080	BF140977.D	12/26/2024	19:25