

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN122724

Review By	John Carlone	Review On	12/30/2024 9:27:13 AM
Supervise By	Maresh Dadoda	Supervise On	12/30/2024 12:45:11 PM
SubDirectory	VN122724	HP Acquire Method	HP Processing Method 82N122724W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP132335		
Initial Calibration Stds	VP132354,VP132355,VP132356,VP132357,VP132358,VP132359		
CCC	VP132336		
Internal Standard/PEM			
ICV/I.BLK	VP132360		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN085320.D	27 Dec 2024 09:52		JC\MD	Ok
2	VSTDICC100	VSTDICC100	VN085321.D	27 Dec 2024 10:26		JC\MD	Ok,M
3	VSTDICCC050	VSTDICCC050	VN085322.D	27 Dec 2024 10:50		JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN085323.D	27 Dec 2024 11:13		JC\MD	Ok,M
5	VSTDICC010	VSTDICC010	VN085324.D	27 Dec 2024 11:37		JC\MD	Ok,M
6	VSTDICC005	VSTDICC005	VN085325.D	27 Dec 2024 12:01		JC\MD	Ok,M
7	VSTDICC001	VSTDICC001	VN085326.D	27 Dec 2024 12:48	Comp #86 on QR	JC\MD	Ok,M
8	IBLK	IBLK	VN085327.D	27 Dec 2024 13:12		JC\MD	Ok
9	VSTDICV050	ICVVN122724	VN085328.D	27 Dec 2024 15:58		JC\MD	Ok,M
10	VN1227WBL01	VN1227WBL01	VN085329.D	27 Dec 2024 16:33		JC\MD	Ok
11	VN1227WBS01	VN1227WBS01	VN085330.D	27 Dec 2024 16:57		JC\MD	Ok,M
12	VN1227WBSD01	VN1227WBSD01	VN085331.D	27 Dec 2024 17:30		JC\MD	Ok,M
13	PB165857TB	PB165857TB	VN085332.D	27 Dec 2024 17:55		JC\MD	Ok
14	P5362-02	WC-SOIL-20241219	VN085333.D	27 Dec 2024 18:18	vial A pH#5.0	JC\MD	Ok
15	P5380-02	TAPIAL3-IDW-SOIL-12	VN085334.D	27 Dec 2024 18:42	vial A pH#5.0	JC\MD	Ok
16	P5386-02	MOO-24-00398	VN085335.D	27 Dec 2024 19:06	vial A pH#5.0	JC\MD	Ok,M
17	P5386-04	MOO-24-00395-96	VN085336.D	27 Dec 2024 19:30	vial A pH#5.0	JC\MD	Ok,M
18	VSTDCCC050	VSTDCCC050EC	VN085337.D	27 Dec 2024 19:54		JC\MD	Ok,M

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN122724

Review By	John Carlone	Review On	12/30/2024 9:27:13 AM
Supervise By	Mahesh Dadoda	Supervise On	12/30/2024 12:45:11 PM
SubDirectory	VN122724	HP Acquire Method	HP Processing Method 82N122724W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP132335		
Initial Calibration Stds	VP132354,VP132355,VP132356,VP132357,VP132358,VP132359		
CCC	VP132336		
Internal Standard/PEM			
ICV/I.BLK	VP132360		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

M : Manual Integration