

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5362

Client: PARSONS Main of New York, Inc.

Analytical Method: SW8260D

Datafile : VN085330.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1227WBS01	Vinyl chloride	20	16.8	ug/L	84			65	117	
	1,1-Dichloroethene	20	17.4	ug/L	87			74	110	
	2-Butanone	100	96.2	ug/L	96			65	122	
	Carbon Tetrachloride	20	17.6	ug/L	88			77	113	
	Chloroform	20	17.4	ug/L	87			79	113	
	Benzene	20	17.5	ug/L	88			82	109	
	1,2-Dichloroethane	20	17.3	ug/L	86			80	115	
	Trichloroethene	20	16.8	ug/L	84			77	113	
	Tetrachloroethene	20	18.0	ug/L	90			67	123	
	Chlorobenzene	20	17.7	ug/L	89			82	109	