

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5362

Client: PARSONS Main of New York, Inc.

Analytical Method: SW8260D

Datafile : VN085331.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1227WBSD01	Vinyl chloride	20	18.6	ug/L	93	10		65	117	20
	1,1-Dichloroethene	20	18.6	ug/L	93	7		74	110	20
	2-Butanone	100	110	ug/L	110	14		65	122	20
	Carbon Tetrachloride	20	19.3	ug/L	97	10		77	113	20
	Chloroform	20	19.1	ug/L	96	10		79	113	20
	Benzene	20	19.5	ug/L	98	11		82	109	20
	1,2-Dichloroethane	20	19.7	ug/L	99	14		80	115	20
	Trichloroethene	20	18.5	ug/L	93	10		77	113	20
	Tetrachloroethene	20	18.9	ug/L	95	5		67	123	20
	Chlorobenzene	20	18.6	ug/L	93	4		82	109	20