

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
 Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362
 Lab File ID: VN085322.D Date Analyzed: 12/27/2024
 Instrument ID: MSVOA_N Time Analyzed: 10:50
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	113462	13.788				
UPPER LIMIT	226924	14.288				
LOWER LIMIT	56731	13.288				
EPA SAMPLE NO.						
WC-SOIL-20241219	92734	13.79				
VN1227WBL01	90060	13.79				
VN1227WBS01	111138	13.79				
VN1227WBSD01	101437	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.