

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY122624

Review By	Mahesh Dadoda	Review On	12/27/2024 9:56:45 AM
Supervise By	Semsettin Yesilyurt	Supervise On	12/27/2024 10:00:03 AM
SubDirectory	VY122624	HP Acquire Method	MSVOA_Y
		HP Processing Method	82Y122624S.M

STD. NAME	STD REF.#
Tune/Reschk	VP132321
Initial Calibration Stds	VP132326,VP132327,VP132328,VP132329,VP132330,VP132331
CCC	VP132333,VP132334
Internal Standard/PEM	
ICV/I.BLK	VP132332
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020705.D	26 Dec 2024 08:29		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY020706.D	26 Dec 2024 09:31	Fail for acroline	SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY020707.D	26 Dec 2024 09:53	LR-11,50	SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY020708.D	26 Dec 2024 10:16		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY020709.D	26 Dec 2024 10:39		SY/MD	Ok,M
6	VSTDICC150	VSTDICC150	VY020710.D	26 Dec 2024 11:01		SY/MD	Ok,M
7	VSTDICC100	VSTDICC100	VY020711.D	26 Dec 2024 11:51		SY/MD	Ok,M
8	VIBLK	VIBLK	VY020712.D	26 Dec 2024 12:13		SY/MD	Ok
9	VSTDICV050	ICVVY122624	VY020713.D	26 Dec 2024 12:51		SY/MD	Ok,M
10	VY1226SBL01	VY1226SBL01	VY020714.D	26 Dec 2024 13:56		SY/MD	Ok
11	VY1226SBS01	VY1226SBS01	VY020715.D	26 Dec 2024 14:32		SY/MD	Ok,M
12	VY1226SBSD01	VY1226SBSD01	VY020716.D	26 Dec 2024 14:54		SY/MD	Ok,M
13	P5383-01	OK-02-12232024	VY020717.D	26 Dec 2024 15:18	Not purged,wrong purge	SY/MD	Not Ok
14	P5316-01	TT-304-IDWSO-202412	VY020718.D	26 Dec 2024 15:41	vial B	SY/MD	Ok
15	P5383-01	OK-02-12232024	VY020719.D	26 Dec 2024 16:04	ISTD Fail.vial B	SY/MD	ReRun
16	P5316-02	TT-304-IDWSO-202412	VY020720.D	26 Dec 2024 16:28	vial B	SY/MD	Ok
17	P5383-01	OK-02-12232024	VY020721.D	26 Dec 2024 16:51	vial A	SY/MD	Ok
18	P5316-03	TT-304-IDWSO-202412	VY020722.D	26 Dec 2024 17:15	vial B	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY122624

Review By	Mahesh Dadoda	Review On	12/27/2024 9:56:45 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	12/27/2024 10:00:03 AM		
SubDirectory	VY122624	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y122624S.M
STD. NAME		STD REF.#			
Tune/Reschk		VP132321			
Initial Calibration Stds		VP132326,VP132327,VP132328,VP132329,VP132330,VP132331			
CCC		VP132333,VP132334			
Internal Standard/PEM					
ICV/I.BLK		VP132332			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P5342-01RE	CHRT26634RE	VY020723.D	26 Dec 2024 17:38	Internal Standard Fail ;Surrogate fail,vial B	SY/MD	Confirms
20	P5342-03RE	HT2651RE	VY020724.D	26 Dec 2024 18:01	Internal Standard Fail,surrogate fail,vial B	SY/MD	Confirms
21	P5342-05RE	RB21198RE	VY020725.D	26 Dec 2024 18:25	Internal Standard Fail,vial B	SY/MD	Confirms
22	P5318-01	AU-06-122024	VY020726.D	26 Dec 2024 18:48	Internal Standard Fail,vial A	SY/MD	ReRun
23	P5382-01	COMP-1	VY020727.D	26 Dec 2024 19:12	vial A	SY/MD	Ok
24	VSTDCCC050	VSTDCCC050EC	VY020728.D	26 Dec 2024 19:34		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY122724

Review By	Mahesh Dadoda	Review On	12/30/2024 12:44:21 PM
Supervise By	Semsettin Yesilyurt	Supervise On	12/30/2024 12:46:36 PM
SubDirectory	VY122724	HP Acquire Method	MSVOA_Y HP Processing Method 82Y122624S.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP132344
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132345,VP132346

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020729.D	27 Dec 2024 09:05		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY020730.D	27 Dec 2024 10:20		SY/MD	Ok,M
3	VY1227SBL01	VY1227SBL01	VY020731.D	27 Dec 2024 11:24		SY/MD	Ok
4	VY1227SBS01	VY1227SBS01	VY020732.D	27 Dec 2024 12:03		SY/MD	Ok,M
5	VY1227SBSD01	VY1227SBSD01	VY020733.D	27 Dec 2024 12:26		SY/MD	Ok,M
6	P5382-02	COMP-2	VY020734.D	27 Dec 2024 12:55	vial A	SY/MD	Ok
7	P5382-03	COMP-3	VY020735.D	27 Dec 2024 13:18	vial A	SY/MD	Ok
8	P5387-01	TR-05-122624	VY020736.D	27 Dec 2024 14:00	NOT PURGE	SY/MD	Not Ok
9	P5318-01	AU-06-122024	VY020737.D	27 Dec 2024 14:24	NOT PURGE	SY/MD	Not Ok
10	P5362-01	WC-SOIL-20241219	VY020738.D	27 Dec 2024 14:47	vial A	SY/MD	Ok
11	P5387-01	TR-05-122624	VY020739.D	27 Dec 2024 15:10	NOT PURGE	SY/MD	Not Ok
12	P5386-01	MOO-24-00398	VY020740.D	27 Dec 2024 15:34	Need MeOH,vial A	SY/MD	Dilution
13	VIBLK	VIBLK	VY020741.D	27 Dec 2024 15:57		SY/MD	Ok
14	VIBLK	VIBLK	VY020742.D	27 Dec 2024 16:21		SY/MD	Ok
15	P5318-01	AU-06-122024	VY020743.D	27 Dec 2024 16:44	Internal standard fail,vial B	SY/MD	Ok
16	VIBLK	VIBLK	VY020744.D	27 Dec 2024 17:07		SY/MD	Ok
17	P5387-01	TR-05-122624	VY020745.D	27 Dec 2024 17:31	vial A	SY/MD	Ok
18	P5390-01	EO-01-12272024	VY020746.D	27 Dec 2024 17:54	Internal Standard Fail,vial A	SY/MD	ReRun

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY122724

Review By	Mahesh Dadoda	Review On	12/30/2024 12:44:21 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	12/30/2024 12:46:36 PM		
SubDirectory	VY122724	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y122624S.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP132344
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132345,VP132346

19	P5391-01	NB-07-12272024	VY020747.D	27 Dec 2024 18:18	NOT PURGE	SY/MD	Not Ok
20	VIBLK	VIBLK	VY020748.D	27 Dec 2024 18:41		SY/MD	Ok
21	VSTDCCC050	VSTDCCC050EC	VY020749.D	27 Dec 2024 19:04		SY/MD	Ok,M

M : Manual Integration