

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
 Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362
 Lab File ID: VY020730.D Date Analyzed: 12/27/2024
 Instrument ID: MSVOA_Y Time Analyzed: 10:20
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	79722	13.358				
UPPER LIMIT	159444	13.858				
LOWER LIMIT	39861	12.858				
EPA SAMPLE NO.						
WC-SOIL-20241219	57023	13.35				
VY1227SBL01	62475	13.35				
VY1227SBS01	85032	13.36				
VY1227SBSD01	79149	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.