

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51821 MA No. : SDG No.: MJNKA0Initial Calibration Verification Source : MP83629Continuing Calibration Verification Source : MP83632Run Batch: LB134112 Analytical Method: ICP-MSConcentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV002				ID: CCV011				ID: CCV012		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic	200	220	109	1	500	490	99	1	490	98	1
Copper	100	100	101	0	5000	4800	96	0	4800	96	1
Lead	200	200	101	1	2500	2600	103	1	2600	103	1
Zinc	200	210	105	1	5000	4700	95	0	4700	95	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51821 MA No. : SDG No.: MJNKA0Initial Calibration Verification Source : MP83629Continuing Calibration Verification Source : MP83632Run Batch: LB134112 Analytical Method: ICP-MSConcentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV013				ID: CCV014		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic					500	490	99	1	490	99	0
Copper					5000	4800	97	0	4900	97	0
Lead					2500	2600	102	1	2600	102	2
Zinc					5000	4700	95	1	4800	96	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51821 MA No. : SDG No.: MJNKA0
 Initial Calibration Verification Source : MP83629
 Continuing Calibration Verification Source : MP83632
 Run Batch: LB134112 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV015				ID: CCV016		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic					500	490	99	2	500	99	0
Copper					5000	4900	99	0	4900	99	1
Lead					2500	2600	104	0	2500	102	3
Zinc					5000	4900	97	1	4800	97	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51821 MA No. : SDG No.: MJNKA0
 Initial Calibration Verification Source : MP83629
 Continuing Calibration Verification Source : MP83632
 Run Batch: LB134112 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV017				ID: CCV018		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic					500	490	99	1	490	98	1
Copper					5000	4900	98	0	4900	98	0
Lead					2500	2600	103	1	2600	103	2
Zinc					5000	4900	97	1	4800	97	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51821 MA No. : SDG No.: MJNKA0Initial Calibration Verification Source : MP83629Continuing Calibration Verification Source : MP83632Run Batch: LB134112 Analytical Method: ICP-MSConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV019				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic					500	490	97	1			
Copper					5000	4900	98	1			
Lead					2500	2500	101	1			
Zinc					5000	4800	96	1			