

## LABORATORY CONTROL SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51821 MA No.: \_\_\_\_\_ SDG No.: MJNKA0  
Matrix : Water Preparation Method: 200.8  
Analytical Method: ICP-MS Preparation Batch: PB165859  
Concentration Units ( $\mu\text{g/L}$ ,  $\text{mg/L}$ ,  $\text{mg/kg}$  dry weight, or  $\mu\text{g}$ ):  $\mu\text{g/L}$

Analyte	True	Found	%R
Arsenic	2	2.3	113
Copper	4	4.3	108
Lead	2	1.9	97
Zinc	10	11	107

## LABORATORY CONTROL SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51821 MA No.: \_\_\_\_\_ SDG No.: MJNKA0  
Matrix : Soil Preparation Method: 200.8  
Analytical Method: ICP-MS Preparation Batch: PB165866  
Concentration Units ( $\mu\text{g/L}$ ,  $\text{mg/L}$ ,  $\text{mg/kg}$  dry weight, or  $\mu\text{g}$ ): mg/kg

Analyte	True	Found	%R
Arsenic	1	1.1	114
Copper	2	2.2	109
Lead	1	0.99	99
Zinc	5	5.5	110