

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134184

Review By	Sarabjit Jaswal	Review On	1/8/2025 8:28:15 AM
Supervise By	Mohan Bera	Supervise On	1/8/2025 8:29:11 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard			
Chk Standard	MP83635,MP83634		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE001	TUNE	01/07/25 12:04		Jaswal	OK
2	S00	S0	CAL1	01/07/25 12:36		Jaswal	OK
3	S02	S02	CAL3	01/07/25 12:43		Jaswal	OK
4	S03	S03	CAL4	01/07/25 12:46		Jaswal	OK
5	S04	S04	CAL5	01/07/25 12:49		Jaswal	OK
6	S05	S05	CAL6	01/07/25 12:52		Jaswal	OK
7	S06	S06	CAL7	01/07/25 12:55		Jaswal	OK
8	S07	S07	CAL8	01/07/25 12:58		Jaswal	OK
9	S08	S08	CAL9	01/07/25 13:00		Jaswal	OK
10	ICV001	ICV001	ICV	01/07/25 13:15		Jaswal	OK
11	ICB001	ICB001	ICB	01/07/25 13:27	Inst. paused for review	Jaswal	OK
12	ICSA001	ICSA001	ICSA	01/07/25 13:30		Jaswal	OK
13	ICSAB001	ICSAB001	ICSAB	01/07/25 13:39	Inst. paused for review	Jaswal	OK
14	CCV001	CCV001	CCV	01/07/25 13:42		Jaswal	OK
15	CCB001	CCB001	CCB	01/07/25 13:45		Jaswal	OK
16	P5364-06	MJNKB1	SAM	01/07/25 13:48	As,Pb,Zn High	Jaswal	Dilution
17	P5364-07	MJNKB1D	DUP	01/07/25 13:51	As,Pb,Zn High	Jaswal	Dilution
18	P5364-06L	MJNKB1L	SD	01/07/25 13:54	As,Pb,Zn High	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134184

Review By	Sarabjit Jaswal	Review On	1/8/2025 8:28:15 AM
Supervise By	Mohan Bera	Supervise On	1/8/2025 8:29:11 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard			
Chk Standard	MP83635,MP83634		

19	P5364-08	MJNKB1S	MS	01/07/25 13:57	As,Pb,Zn High	Jaswal	Dilution
20	P5364-06DL	MJNKB1	SAM	01/07/25 14:20	2X For As,Pb,Zn, Still As,Zn High	Jaswal	Dilution
21	P5364-07DL	MJNKB1D	DUP	01/07/25 14:22	2X For As,Pb,Zn, Still As,Zn High	Jaswal	Dilution
22	P5364-06LDL	MJNKB1L	SD	01/07/25 14:25	10X For As,Pb,Zn, Still As,Zn High	Jaswal	Dilution
23	P5364-08DL	MJNKB1S	MS	01/07/25 14:28	2X For As,Pb,Zn, Still As,Zn High	Jaswal	Dilution
24	P5364-06DL2	MJNKB1	SAM	01/07/25 14:31	5X For As,Zn,Still As High	Jaswal	Dilution
25	P5364-07DL2	MJNKB1D	DUP	01/07/25 14:34	5X For As,Zn,Still As High	Jaswal	Dilution
26	P5364-06LDL2	MJNKB1L	SD	01/07/25 14:38	25X For As,Zn,Still As High	Jaswal	Dilution
27	P5364-08DL2	MJNKB1S	MS	01/07/25 14:41	5X For As,Zn,Still As High	Jaswal	Dilution
28	P5364-06DL3	MJNKB1	SAM	01/07/25 14:44	20X For As	Jaswal	Confirms
29	P5364-07DL3	MJNKB1D	DUP	01/07/25 14:47	20X For As	Jaswal	Confirms
30	P5364-06LDL3	MJNKB1L	SD	01/07/25 14:50	100X For As	Jaswal	Confirms
31	P5364-08DL3	MJNKB1S	MS	01/07/25 14:53	20X For As	Jaswal	Confirms
32	CCV002	CCV002	CCV	01/07/25 14:56		Jaswal	OK
33	CCB002	CCB002	CCB	01/07/25 14:59		Jaswal	OK
34	P5364-06A	MJNKB1A	PS	01/07/25 15:52	Not Required	Jaswal	Not Ok
35	CCV003	CCV003	CCV	01/07/25 16:30		Jaswal	OK
36	CCB003	CCB003	CCB	01/07/25 16:32		Jaswal	OK