

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134040

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/20/2024	11:40
2	CAL2	1	Water	NA	NA	20.3	7.00	12/20/2024	11:41
3	CAL3	1	Water	NA	NA	20.2	10.02	12/20/2024	11:44
4	ICV	1	Water	NA	NA	20.2	7.01	12/20/2024	11:45
5	CCV1	1	Water	NA	NA	20.3	2.02	12/20/2024	11:47
6	P5365-01	1	Solid	20.02	20	23.6	6.89	12/20/2024	11:55
7	P5365-01DUP	1	Solid	20.03	20	23.7	6.90	12/20/2024	11:56
8	CCV2	1	Water	NA	NA	20.2	12.02	12/20/2024	12:00

WORKLIST(Hardcopy Internal Chain)

WB 134040

WorkList Name : PH P5365

WorkList ID : 186529

Department : Wet-Chemistry

Date : 12-20-2024 11:18:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5365-01	TAPFTA-SB01L-4-5-121924-00	Solid	pH	Cool 4 deg C	WEST04	N21	12/19/2024	9045D

Date/Time 12/20/24 11:30
Raw Sample Received by: ad ccc
Raw Sample Relinquished by: ad sm

Date/Time 12/20/24 13:13
Raw Sample Received by: ad sm
Raw Sample Relinquished by: ad ccc