

Laboratory Control Sample Summary

Client: Weston Solutions

SDG No.: P5365

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Run No.: LB134042

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134042BSS							
Bromide	mg/Kg	200	211		106	1	90-110	12/20/2024
Chloride	mg/Kg	60	62.9		105	1	90-110	12/20/2024
Fluoride	mg/Kg	40	42.6		106	1	90-110	12/20/2024
Nitrite	mg/Kg	60	62.9		105	1	90-110	12/20/2024
Nitrate	mg/Kg	50	52.6		105	1	90-110	12/20/2024
Sulfate	mg/Kg	300	314		105	1	90-110	12/20/2024
Orthophosphate as P	mg/Kg	100	109		109	1	90-110	12/20/2024

Laboratory Control Sample Summary

Client: Weston Solutions

SDG No.: P5365

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Run No.: LB134163

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134163BSS							
TOC	mg/Kg	1000	1020		102	1	90-110	01/06/2025