



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Weston Solutions **SDG No.:** P5365  
**Contract:** WEST04 **Lab Code:** CHEM **Case No.:** P5365 **SAS No.:** P5365

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | LOD  | CRQL | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|------|----|------------------|------------------|---------------|
| ICB28     | Mercury | 0.20           | +/-0.20             | U            | 0.16 | 0.20 | CV | 12/20/2024       | 14:15            | LB134050      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

| <b>Client:</b> <u>Weston Solutions</u> |         | <b>SDG No.:</b> <u>P5365</u> |                     |                               |      |      |                              |                  |                  |               |
|----------------------------------------|---------|------------------------------|---------------------|-------------------------------|------|------|------------------------------|------------------|------------------|---------------|
| <b>Contract:</b> <u>WEST04</u>         |         | <b>Lab Code:</b> <u>CHEM</u> |                     | <b>Case No.:</b> <u>P5365</u> |      |      | <b>SAS No.:</b> <u>P5365</u> |                  |                  |               |
| Sample ID                              | Analyte | Result<br>ug/L               | Acceptance<br>Limit | Conc<br>Qual                  | LOD  | CRQL | M                            | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB57                                  | Mercury | 0.20                         | +/-0.20             | U                             | 0.16 | 0.20 | CV                           | 12/20/2024       | 14:20            | LB134050      |
| CCB58                                  | Mercury | 0.20                         | +/-0.20             | U                             | 0.16 | 0.20 | CV                           | 12/20/2024       | 14:45            | LB134050      |
| CCB59                                  | Mercury | 0.20                         | +/-0.20             | U                             | 0.16 | 0.20 | CV                           | 12/20/2024       | 15:03            | LB134050      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions SDG No.: P5365  
Contract: WEST04 Lab Code: CHEM Case No.: P5365 SAS No.: P5365

| Sample ID | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | LOD  | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|---------------------|--------------|------|------|---|------------------|------------------|---------------|
| ICB01     | Aluminum  | 20.0           | +/-20.0             | U            | 10.0 | 20.0 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Antimony  | 2.00           | +/-2.00             | U            | 0.25 | 2.00 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Arsenic   | 1.00           | +/-1.00             | U            | 0.25 | 1.00 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Barium    | 10.0           | +/-10.0             | U            | 1.25 | 10.0 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Beryllium | 1.00           | +/-1.00             | U            | 0.25 | 1.00 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Cadmium   | 1.00           | +/-1.00             | U            | 0.50 | 1.00 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Calcium   | 500            | +/-500              | U            | 190  | 500  | P | 01/06/2025       | 14:24            | LB134187      |
|           | Chromium  | 2.00           | +/-2.00             | U            | 0.75 | 2.00 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Cobalt    | 1.00           | +/-1.00             | U            | 0.25 | 1.00 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Copper    | 2.00           | +/-2.00             | U            | 1.50 | 2.00 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Iron      | 50.0           | +/-50.0             | U            | 25.0 | 50.0 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Lead      | 1.00           | +/-1.00             | U            | 0.75 | 1.00 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Magnesium | 500            | +/-500              | U            | 190  | 500  | P | 01/06/2025       | 14:24            | LB134187      |
|           | Manganese | 1.00           | +/-1.00             | U            | 0.75 | 1.00 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Nickel    | 1.00           | +/-1.00             | U            | 0.25 | 1.00 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Potassium | 500            | +/-500              | U            | 190  | 500  | P | 01/06/2025       | 14:24            | LB134187      |
|           | Selenium  | 5.00           | +/-5.00             | U            | 4.50 | 5.00 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Silver    | 1.00           | +/-1.00             | U            | 0.50 | 1.00 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Sodium    | 500            | +/-500              | U            | 190  | 500  | P | 01/06/2025       | 14:24            | LB134187      |
|           | Thallium  | 1.00           | +/-1.00             | U            | 0.50 | 1.00 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Vanadium  | 5.00           | +/-5.00             | U            | 0.25 | 5.00 | P | 01/06/2025       | 14:24            | LB134187      |
|           | Zinc      | 5.00           | +/-5.00             | U            | 1.50 | 5.00 | P | 01/06/2025       | 14:24            | LB134187      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

| <b>Client:</b> <u>Weston Solutions</u> |                              | <b>SDG No.:</b> <u>P5365</u>  |                              |              |      |      |   |                  |                  |               |
|----------------------------------------|------------------------------|-------------------------------|------------------------------|--------------|------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> <u>WEST04</u>         | <b>Lab Code:</b> <u>CHEM</u> | <b>Case No.:</b> <u>P5365</u> | <b>SAS No.:</b> <u>P5365</u> |              |      |      |   |                  |                  |               |
| Sample ID                              | Analyte                      | Result<br>ug/L                | Acceptance<br>Limit          | Conc<br>Qual | LOD  | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| <b>CCB01</b>                           | Aluminum                     | 20.0                          | +/-20.0                      | U            | 10.0 | 20.0 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Antimony                     | 2.00                          | +/-2.00                      | U            | 0.25 | 2.00 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Arsenic                      | 1.00                          | +/-1.00                      | U            | 0.25 | 1.00 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Barium                       | 10.0                          | +/-10.0                      | U            | 1.25 | 10.0 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Beryllium                    | 1.00                          | +/-1.00                      | U            | 0.25 | 1.00 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Cadmium                      | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Calcium                      | 500                           | +/-500                       | U            | 190  | 500  | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Chromium                     | 2.00                          | +/-2.00                      | U            | 0.75 | 2.00 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Cobalt                       | 1.00                          | +/-1.00                      | U            | 0.25 | 1.00 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Copper                       | 2.00                          | +/-2.00                      | U            | 1.50 | 2.00 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Iron                         | 50.0                          | +/-50.0                      | U            | 25.0 | 50.0 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Lead                         | 1.00                          | +/-1.00                      | U            | 0.75 | 1.00 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Magnesium                    | 500                           | +/-500                       | U            | 190  | 500  | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Manganese                    | 1.00                          | +/-1.00                      | U            | 0.75 | 1.00 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Nickel                       | 1.00                          | +/-1.00                      | U            | 0.25 | 1.00 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Potassium                    | 500                           | +/-500                       | U            | 190  | 500  | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Selenium                     | 5.00                          | +/-5.00                      | U            | 4.50 | 5.00 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Silver                       | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Sodium                       | 500                           | +/-500                       | U            | 190  | 500  | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Thallium                     | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Vanadium                     | 5.00                          | +/-5.00                      | U            | 0.25 | 5.00 | P | 01/06/2025       | 14:41            | LB134187      |
|                                        | Zinc                         | 5.00                          | +/-5.00                      | U            | 1.50 | 5.00 | P | 01/06/2025       | 14:41            | LB134187      |
| <b>CCB02</b>                           | Aluminum                     | 3.09                          | +/-20.0                      | J            | 10.0 | 20.0 | P | 01/06/2025       | 15:38            | LB134187      |
|                                        | Antimony                     | 0.13                          | +/-2.00                      | J            | 0.25 | 2.00 | P | 01/06/2025       | 15:38            | LB134187      |
|                                        | Arsenic                      | 1.00                          | +/-1.00                      | U            | 0.25 | 1.00 | P | 01/06/2025       | 15:38            | LB134187      |
|                                        | Barium                       | 10.0                          | +/-10.0                      | U            | 1.25 | 10.0 | P | 01/06/2025       | 15:38            | LB134187      |
|                                        | Beryllium                    | 1.00                          | +/-1.00                      | U            | 0.25 | 1.00 | P | 01/06/2025       | 15:38            | LB134187      |
|                                        | Cadmium                      | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 01/06/2025       | 15:38            | LB134187      |
|                                        | Calcium                      | 500                           | +/-500                       | U            | 190  | 500  | P | 01/06/2025       | 15:38            | LB134187      |
|                                        | Chromium                     | 2.00                          | +/-2.00                      | U            | 0.75 | 2.00 | P | 01/06/2025       | 15:38            | LB134187      |
|                                        | Cobalt                       | 1.00                          | +/-1.00                      | U            | 0.25 | 1.00 | P | 01/06/2025       | 15:38            | LB134187      |
|                                        | Copper                       | 2.00                          | +/-2.00                      | U            | 1.50 | 2.00 | P | 01/06/2025       | 15:38            | LB134187      |
|                                        | Iron                         | 50.0                          | +/-50.0                      | U            | 25.0 | 50.0 | P | 01/06/2025       | 15:38            | LB134187      |
|                                        | Lead                         | 0.17                          | +/-1.00                      | J            | 0.75 | 1.00 | P | 01/06/2025       | 15:38            | LB134187      |
|                                        | Magnesium                    | 500                           | +/-500                       | U            | 190  | 500  | P | 01/06/2025       | 15:38            | LB134187      |
|                                        | Manganese                    | 0.32                          | +/-1.00                      | J            | 0.75 | 1.00 | P | 01/06/2025       | 15:38            | LB134187      |
|                                        | Nickel                       | 1.00                          | +/-1.00                      | U            | 0.25 | 1.00 | P | 01/06/2025       | 15:38            | LB134187      |
|                                        | Potassium                    | 500                           | +/-500                       | U            | 190  | 500  | P | 01/06/2025       | 15:38            | LB134187      |
|                                        | Selenium                     | 5.00                          | +/-5.00                      | U            | 4.50 | 5.00 | P | 01/06/2025       | 15:38            | LB134187      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Weston Solutions **SDG No.:** P5365  
**Contract:** WEST04 **Lab Code:** CHEM **Case No.:** P5365 **SAS No.:** P5365

| Sample ID | Analyte  | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | LOD  | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|---------------------|--------------|------|------|---|------------------|------------------|---------------|
| CCB02     | Silver   | 1.00           | +/-1.00             | U            | 0.50 | 1.00 | P | 01/06/2025       | 15:38            | LB134187      |
|           | Sodium   | 500            | +/-500              | U            | 190  | 500  | P | 01/06/2025       | 15:38            | LB134187      |
|           | Thallium | 1.00           | +/-1.00             | U            | 0.50 | 1.00 | P | 01/06/2025       | 15:38            | LB134187      |
|           | Vanadium | 5.00           | +/-5.00             | U            | 0.25 | 5.00 | P | 01/06/2025       | 15:38            | LB134187      |
|           | Zinc     | 5.00           | +/-5.00             | U            | 1.50 | 5.00 | P | 01/06/2025       | 15:38            | LB134187      |