

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions **SDG No.:** P5365
Contract: WEST04 **Lab Code:** CHEM **Case No.:** P5365 **SAS No.:** P5365

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165798BS Mercury	mg/Kg	0.25	0.26		106	80 - 124	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions **SDG No.:** P5365
Contract: WEST04 **Lab Code:** CHEM **Case No.:** P5365 **SAS No.:** P5365

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165957BS							
Aluminum	mg/Kg	780	772		99	78 - 124	P
Antimony	mg/Kg	38.8	41.2		106	72 - 124	P
Arsenic	mg/Kg	38.8	39.1		101	82 - 118	P
Barium	mg/Kg	190	207		109	86 - 116	P
Beryllium	mg/Kg	38.8	41.9		108	80 - 120	P
Cadmium	mg/Kg	38.8	40.6		105	84 - 116	P
Calcium	mg/Kg	3900	4230		108	86 - 118	P
Chromium	mg/Kg	38.8	39.1		101	83 - 119	P
Cobalt	mg/Kg	38.8	39.9		103	84 - 115	P
Copper	mg/Kg	390	398		102	84 - 119	P
Iron	mg/Kg	1900	2070		109	81 - 124	P
Lead	mg/Kg	190	207		109	84 - 118	P
Magnesium	mg/Kg	3900	3900		100	80 - 123	P
Manganese	mg/Kg	390	390		100	85 - 116	P
Nickel	mg/Kg	38.8	39.6		102	84 - 119	P
Potassium	mg/Kg	1900	1910		100	85 - 119	P
Selenium	mg/Kg	38.8	41.8		108	80 - 119	P
Silver	mg/Kg	38.8	42.4		109	83 - 118	P
Sodium	mg/Kg	3900	4000		103	79 - 125	P
Thallium	mg/Kg	38.8	41.3		106	83 - 118	P
Vanadium	mg/Kg	38.8	39.2		101	82 - 116	P
Zinc	mg/Kg	390	398		102	82 - 119	P