

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Weston Solutions

Contract: WEST04

Lab Code: CHEM Case no.: P5365

Sas No.: P5365 SDG No.: P5365

Instrument ID: P7

Start Date : 01/06/2025

Run Number: LB134187

End Date : 01/06/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Non-Collision Cell |   |                 |   |                |   |                  |   |                  |   |
|-----------------|--------------------|------|-----------------------------------------------|---|-----------------|---|----------------|---|------------------|---|------------------|---|
|                 |                    |      | Element<br>6Li                                | Q | Element<br>45Sc | Q | Element<br>89Y | Q | Element<br>103Rh | Q | Element<br>159Tb | Q |
| S0              | S0                 | 1315 | 100                                           |   | 100             |   | 100            |   | 100              |   | 100              |   |
| S2              | S2                 | 1319 | 102                                           |   | 100             |   | 100            |   | 99               |   | 100              |   |
| S3              | S3                 | 1325 | 106                                           |   | 100             |   | 99             |   | 98               |   | 101              |   |
| S4              | S4                 | 1328 | 109                                           |   | 96              |   | 97             |   | 96               |   | 101              |   |
| S5              | S5                 | 1331 | 111                                           |   | 92              |   | 95             |   | 92               |   | 101              |   |
| S6              | S6                 | 1334 | 111                                           |   | 88              |   | 91             |   | 88               |   | 100              |   |
| S7              | S7                 | 1336 | 110                                           |   | 87              |   | 91             |   | 86               |   | 98               |   |
| S8              | S8                 | 1339 | 114                                           |   | 91              |   | 89             |   | 79               |   | 91               |   |
| ICV01           | ICV01              | 1405 | 116                                           |   | 98              |   | 100            |   | 99               |   | 104              |   |
| LLICV01         | LLICV01            | 1409 | 112                                           |   | 97              |   | 98             |   | 97               |   | 101              |   |
| ICB01           | ICB01              | 1424 | 112                                           |   | 99              |   | 99             |   | 98               |   | 102              |   |
| ICSA01          | ICSA01             | 1427 | 113                                           |   | 92              |   | 96             |   | 89               |   | 103              |   |
| ICSAB01         | ICSAB01            | 1431 | 111                                           |   | 93              |   | 97             |   | 91               |   | 106              |   |
| CCV01           | CCV01              | 1434 | 109                                           |   | 90              |   | 94             |   | 86               |   | 101              |   |
| CCB01           | CCB01              | 1441 | 113                                           |   | 95              |   | 98             |   | 97               |   | 104              |   |
| CRI             | CRI                | 1454 | 112                                           |   | 100             |   | 100            |   | 100              |   | 106              |   |
| PB165957BL      | PB165957BL         | 1458 | 110                                           |   | 99              |   | 100            |   | 99               |   | 105              |   |
| PB165957BS      | PB165957BS         | 1501 | 105                                           |   | 88              |   | 91             |   | 87               |   | 98               |   |
| P5365-01        | TAPFTA-SB01        | 1503 | 117                                           |   | 95              |   | 99             |   | 96               |   | 106              |   |
| P5365-01DUP     | TAPFTA-SB01        | 1507 | 116                                           |   | 96              |   | 100            |   | 96               |   | 105              |   |
| P5365-01L       | TAPFTA-SB01        | 1510 | 114                                           |   | 98              |   | 101            |   | 100              |   | 108              |   |
| P5365-01MS      | TAPFTA-SB01        | 1513 | 115                                           |   | 98              |   | 103            |   | 98               |   | 107              |   |
| P5365-01MSD     | TAPFTA-SB01        | 1516 | 115                                           |   | 99              |   | 103            |   | 99               |   | 107              |   |
| P5365-01A       | TAPFTA-SB01        | 1519 | 113                                           |   | 99              |   | 102            |   | 98               |   | 106              |   |
| CCV02           | CCV02              | 1535 | 109                                           |   | 90              |   | 92             |   | 84               |   | 100              |   |
| CCB02           | CCB02              | 1538 | 118                                           |   | 93              |   | 97             |   | 96               |   | 105              |   |

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Weston Solutions

Contract: WEST04

Lab Code: CHEM Case no.: P5365

Sas No.: P5365 SDG No.: P5365

Instrument ID: P7

Start Date : 01/06/2025

Run Number: LB134187

End Date : 01/06/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Collision Cell |   |                |   |                  |   |                  |   |                  |   |
|-----------------|--------------------|------|-------------------------------------------|---|----------------|---|------------------|---|------------------|---|------------------|---|
|                 |                    |      | Element<br>45Sc                           | Q | Element<br>89Y | Q | Element<br>103Rh | Q | Element<br>159Tb | Q | Element<br>165Ho | Q |
| S0              | S0                 | 1315 | 100                                       |   | 100            |   | 100              |   | 100              |   | 100              |   |
| S2              | S2                 | 1319 | 100                                       |   | 99             |   | 100              |   | 101              |   | 100              |   |
| S3              | S3                 | 1325 | 95                                        |   | 97             |   | 97               |   | 100              |   | 100              |   |
| S4              | S4                 | 1328 | 90                                        |   | 93             |   | 92               |   | 99               |   | 99               |   |
| S5              | S5                 | 1331 | 85                                        |   | 89             |   | 89               |   | 98               |   | 99               |   |
| S6              | S6                 | 1334 | 81                                        |   | 86             |   | 86               |   | 97               |   | 98               |   |
| S7              | S7                 | 1336 | 81                                        |   | 85             |   | 84               |   | 98               |   | 98               |   |
| S8              | S8                 | 1339 | 86                                        |   | 86             |   | 79               |   | 93               |   | 94               |   |
| ICV01           | ICV01              | 1405 | 92                                        |   | 95             |   | 96               |   | 101              |   | 100              |   |
| LLICV01         | LLICV01            | 1409 | 93                                        |   | 96             |   | 97               |   | 100              |   | 101              |   |
| ICB01           | ICB01              | 1424 | 96                                        |   | 98             |   | 99               |   | 101              |   | 101              |   |
| ICSA01          | ICSA01             | 1427 | 85                                        |   | 90             |   | 88               |   | 102              |   | 102              |   |
| ICSAB01         | ICSAB01            | 1431 | 85                                        |   | 90             |   | 89               |   | 101              |   | 102              |   |
| CCV01           | CCV01              | 1434 | 85                                        |   | 88             |   | 84               |   | 99               |   | 99               |   |
| CCB01           | CCB01              | 1441 | 89                                        |   | 92             |   | 96               |   | 101              |   | 102              |   |
| CRI             | CRI                | 1454 | 94                                        |   | 97             |   | 99               |   | 102              |   | 102              |   |
| PB165957BL      | PB165957BL         | 1458 | 94                                        |   | 97             |   | 99               |   | 102              |   | 103              |   |
| PB165957BS      | PB165957BS         | 1501 | 83                                        |   | 87             |   | 88               |   | 99               |   | 101              |   |
| P5365-01        | TAPFTA-SB01        | 1503 | 90                                        |   | 95             |   | 96               |   | 102              |   | 104              |   |
| P5365-01DUP     | TAPFTA-SB01        | 1507 | 91                                        |   | 96             |   | 97               |   | 103              |   | 104              |   |
| P5365-01L       | TAPFTA-SB01        | 1510 | 91                                        |   | 96             |   | 98               |   | 103              |   | 104              |   |
| P5365-01MS      | TAPFTA-SB01        | 1513 | 93                                        |   | 97             |   | 96               |   | 104              |   | 104              |   |
| P5365-01MSD     | TAPFTA-SB01        | 1516 | 92                                        |   | 96             |   | 96               |   | 103              |   | 104              |   |
| P5365-01A       | TAPFTA-SB01        | 1519 | 91                                        |   | 96             |   | 96               |   | 104              |   | 104              |   |
| CCV02           | CCV02              | 1535 | 83                                        |   | 87             |   | 83               |   | 98               |   | 99               |   |
| CCB02           | CCB02              | 1538 | 87                                        |   | 92             |   | 95               |   | 102              |   | 102              |   |