

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG NO.: P5369

Continuing Calib Date: 12/23/2024 Initial Calibration Date(s): 12/20/2024 12/20/2024

Continuing Calib Time: 17:35 Initial Calibration Time(s): 11:38 21:11

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND       |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------|-----|------------|-----------|------------------------|------|------------|
| Aroclor-1016-1 | (1) | 4.85       | 4.85      | 4.78                   | 4.92 | 0.00       |
| Aroclor-1016-2 | (2) | 4.87       | 4.87      | 4.80                   | 4.94 | 0.00       |
| Aroclor-1016-3 | (3) | 4.94       | 4.94      | 4.87                   | 5.01 | 0.00       |
| Aroclor-1016-4 | (4) | 5.04       | 5.04      | 4.97                   | 5.11 | 0.00       |
| Aroclor-1016-5 | (5) | 5.35       | 5.35      | 5.28                   | 5.42 | 0.00       |
| Aroclor-1260-1 | (1) | 6.57       | 6.57      | 6.50                   | 6.64 | 0.00       |
| Aroclor-1260-2 | (2) | 6.85       | 6.85      | 6.78                   | 6.92 | 0.00       |
| Aroclor-1260-3 | (3) | 7.24       | 7.24      | 7.17                   | 7.31 | 0.00       |
| Aroclor-1260-4 | (4) | 7.48       | 7.48      | 7.41                   | 7.55 | 0.00       |
| Aroclor-1260-5 | (5) | 7.82       | 7.82      | 7.75                   | 7.89 | 0.00       |
| TCX            |     | 3.63       | 3.63      | 3.58                   | 3.68 | 0.00       |
| DCB            |     | 9.52       | 9.52      | 9.42                   | 9.62 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG NO.: P5369

Continuing Calib Date: 12/23/2024 Initial Calibration Date(s): 12/20/2024 12/20/2024

Continuing Calib Time: 17:35 Initial Calibration Time(s): 11:38 21:11

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND       |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------|-----|------------|-----------|-----------|------|------------|
|                |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1 | (1) | 4.08       | 4.08      | 4.01      | 4.15 | 0.00       |
| Aroclor-1016-2 | (2) | 4.10       | 4.10      | 4.03      | 4.17 | 0.00       |
| Aroclor-1016-3 | (3) | 4.28       | 4.28      | 4.21      | 4.35 | 0.00       |
| Aroclor-1016-4 | (4) | 4.33       | 4.33      | 4.26      | 4.40 | 0.00       |
| Aroclor-1016-5 | (5) | 4.55       | 4.55      | 4.48      | 4.62 | 0.00       |
| Aroclor-1260-1 | (1) | 5.65       | 5.65      | 5.58      | 5.72 | 0.01       |
| Aroclor-1260-2 | (2) | 5.85       | 5.85      | 5.78      | 5.92 | 0.00       |
| Aroclor-1260-3 | (3) | 6.01       | 6.01      | 5.94      | 6.08 | 0.00       |
| Aroclor-1260-4 | (4) | 6.52       | 6.52      | 6.45      | 6.59 | 0.00       |
| Aroclor-1260-5 | (5) | 6.78       | 6.78      | 6.71      | 6.85 | 0.00       |
| TCX            |     | 2.96       | 2.96      | 2.91      | 3.01 | 0.00       |
| DCB            |     | 8.24       | 8.24      | 8.14      | 8.34 | 0.00       |