

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG NO.: P5369

Continuing Calib Date: 12/23/2024 Initial Calibration Date(s): 12/20/2024 12/20/2024

Continuing Calib Time: 18:04 Initial Calibration Time(s): 11:38 21:11

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.85	4.85	4.78	4.92	0.00
Aroclor-1242-2	(2)	4.87	4.87	4.80	4.94	0.00
Aroclor-1242-3	(3)	4.94	4.94	4.87	5.01	0.00
Aroclor-1242-4	(4)	5.04	5.04	4.97	5.11	0.00
Aroclor-1242-5	(5)	5.83	5.83	5.76	5.90	0.01
TCX		3.63	3.63	3.58	3.68	0.00
DCB		9.52	9.52	9.42	9.62	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG NO.: P5369

Continuing Calib Date: 12/23/2024 Initial Calibration Date(s): 12/20/2024 12/20/2024

Continuing Calib Time: 18:04 Initial Calibration Time(s): 11:38 21:11

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.08	4.08	4.01	4.15	0.00
Aroclor-1242-2	(2)	4.10	4.10	4.03	4.17	0.00
Aroclor-1242-3	(3)	4.28	4.28	4.21	4.35	0.00
Aroclor-1242-4	(4)	4.37	4.37	4.30	4.44	0.00
Aroclor-1242-5	(5)	4.92	4.92	4.85	4.99	0.00
TCX		2.96	2.96	2.91	3.01	0.00
DCB		8.24	8.24	8.14	8.34	0.00