

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG NO.: P5369

Continuing Calib Date: 12/23/2024 Initial Calibration Date(s): 12/20/2024 12/20/2024

Continuing Calib Time: 18:33 Initial Calibration Time(s): 11:38 21:11

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.76	5.76	5.69	5.83	0.00
Aroclor-1254-2	(2)	5.99	6.00	5.93	6.07	0.01
Aroclor-1254-3	(3)	6.39	6.39	6.32	6.46	0.00
Aroclor-1254-4	(4)	6.70	6.70	6.63	6.77	0.00
Aroclor-1254-5	(5)	7.15	7.16	7.09	7.23	0.01
TCX		3.63	3.63	3.58	3.68	0.00
DCB		9.52	9.53	9.43	9.63	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG NO.: P5369

Continuing Calib Date: 12/23/2024 Initial Calibration Date(s): 12/20/2024 12/20/2024

Continuing Calib Time: 18:33 Initial Calibration Time(s): 11:38 21:11

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	4.92	4.92	4.85	4.99	0.00
Aroclor-1254-2 (2)	5.08	5.08	5.01	5.15	0.00
Aroclor-1254-3 (3)	5.50	5.51	5.44	5.58	0.01
Aroclor-1254-4 (4)	5.75	5.75	5.68	5.82	0.00
Aroclor-1254-5 (5)	6.20	6.20	6.13	6.27	0.00
TCX	2.96	2.96	2.91	3.01	0.00
DCB	8.24	8.24	8.14	8.34	0.01