

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG NO.: P5369

Continuing Calib Date: 12/24/2024 Initial Calibration Date(s): 12/20/2024 12/20/2024

Continuing Calib Time: 06:18 Initial Calibration Time(s): 11:38 21:11

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.85	4.85	4.78	4.92	0.00
Aroclor-1248-2	(2)	5.14	5.14	5.07	5.21	0.00
Aroclor-1248-3	(3)	5.35	5.35	5.28	5.42	0.00
Aroclor-1248-4	(4)	5.79	5.79	5.72	5.86	0.01
Aroclor-1248-5	(5)	5.83	5.83	5.76	5.90	0.01
TCX		3.63	3.63	3.58	3.68	0.00
DCB		9.52	9.52	9.42	9.62	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG NO.: P5369

Continuing Calib Date: 12/24/2024 Initial Calibration Date(s): 12/20/2024 12/20/2024

Continuing Calib Time: 06:18 Initial Calibration Time(s): 11:38 21:11

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.08	4.08	4.01	4.15	0.00
Aroclor-1248-2	(2)	4.33	4.33	4.26	4.40	0.00
Aroclor-1248-3	(3)	4.37	4.37	4.30	4.44	0.00
Aroclor-1248-4	(4)	4.55	4.55	4.48	4.62	0.00
Aroclor-1248-5	(5)	4.96	4.96	4.89	5.03	0.00
TCX		2.96	2.96	2.91	3.01	0.00
DCB		8.24	8.24	8.14	8.34	0.01