

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG NO.: P5369

Continuing Calib Date: 12/24/2024 Initial Calibration Date(s): 12/20/2024 12/20/2024

Continuing Calib Time: 06:33 Initial Calibration Time(s): 11:38 21:11

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND       |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------|-----|------------|-----------|-----------|------|------------|
|                |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1 | (1) | 5.76       | 5.76      | 5.69      | 5.83 | 0.00       |
| Aroclor-1254-2 | (2) | 5.99       | 6.00      | 5.93      | 6.07 | 0.01       |
| Aroclor-1254-3 | (3) | 6.39       | 6.39      | 6.32      | 6.46 | 0.00       |
| Aroclor-1254-4 | (4) | 6.70       | 6.70      | 6.63      | 6.77 | 0.00       |
| Aroclor-1254-5 | (5) | 7.15       | 7.16      | 7.09      | 7.23 | 0.01       |
| TCX            |     | 3.63       | 3.63      | 3.58      | 3.68 | 0.00       |
| DCB            |     | 9.52       | 9.53      | 9.43      | 9.63 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG NO.: P5369

Continuing Calib Date: 12/24/2024 Initial Calibration Date(s): 12/20/2024 12/20/2024

Continuing Calib Time: 06:33 Initial Calibration Time(s): 11:38 21:11

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND       |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------|-----|------------|-----------|-----------|------|------------|
|                |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1 | (1) | 4.92       | 4.92      | 4.85      | 4.99 | 0.00       |
| Aroclor-1254-2 | (2) | 5.08       | 5.08      | 5.01      | 5.15 | 0.00       |
| Aroclor-1254-3 | (3) | 5.50       | 5.51      | 5.44      | 5.58 | 0.01       |
| Aroclor-1254-4 | (4) | 5.75       | 5.75      | 5.68      | 5.82 | 0.00       |
| Aroclor-1254-5 | (5) | 6.20       | 6.20      | 6.13      | 6.27 | 0.00       |
| TCX            |     | 2.96       | 2.96      | 2.91      | 3.01 | 0.00       |
| DCB            |     | 8.24       | 8.24      | 8.14      | 8.34 | 0.01       |