

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CHEM02
Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG No.: P5369
Instrument ID: BNA_G Calibration Date(s): 12/11/2024 12/11/2024
Calibration Time(s): 13:23 17:52

LAB FILE ID:		RRFAL1 = BG063651.D			RRFAL2 = BG063652.D		RRFAL3 = BG063653.D	
		RRFAL4 = BG063654.D			RRFAL5 = BG063655.D		RRFAL6 = BG063656.D	
COMPOUND	RRFAL1	RRFAL2	RRFAL3	RRFAL4	RRFAL5	RRFAL6	RRF	% RSD
1,4-Dioxane	0.567	0.593	0.689	0.576	0.596		0.604	8.1
Benzaldehyde		1.337	1.446	1.195	0.907	0.764	1.130	25.4
Hexachloroethane	0.587	0.609	0.627	0.590	0.553		0.593	4.7
Nitrobenzene	0.424	0.448	0.521	0.504	0.488		0.477	8.4
Isophorone	0.875	0.868	0.989	0.918	0.853		0.901	6.1
2-Nitrophenol	0.139	0.151	0.183	0.180	0.170		0.165	11.6
2,4-Dimethylphenol	0.352	0.361	0.404	0.384	0.365		0.373	5.5
Bis(2-Chloroethoxy)methane	0.467	0.476	0.531	0.494	0.469		0.487	5.5
2,4-Dichlorophenol	0.269	0.299	0.342	0.319	0.311		0.308	8.7
Naphthalene	0.957	1.007	1.105	1.008	0.923		1.000	6.9
4-Chloroaniline		0.347	0.423	0.390	0.372	0.341	0.374	9.0
Hexachlorobutadiene	0.234	0.242	0.271	0.253	0.250		0.250	5.5
Phenol		1.813	2.129	2.027	1.934	1.913	1.963	6.1
Caprolactam		0.099	0.119	0.110	0.101	0.104	0.106	7.7
4-Chloro-3-methylphenol	0.342	0.349	0.383	0.376	0.358		0.362	4.8
2-Methylnaphthalene	0.729	0.703	0.783	0.716	0.662		0.719	6.1
Hexachlorocyclopentadiene		0.151	0.230	0.265	0.297	0.322	0.253	26.4
2,4,6-Trichlorophenol	0.328	0.348	0.392	0.373	0.368		0.362	6.8
2,4,5-Trichlorophenol	0.345	0.352	0.398	0.410	0.384		0.378	7.5
1,1-Biphenyl	1.269	1.316	1.449	1.343	1.217		1.319	6.6
2-Chloronaphthalene	1.014	1.040	1.167	1.084	0.992		1.059	6.5
2-Nitroaniline	0.306	0.319	0.371	0.391	0.387		0.355	11.1
Bis(2-Chloroethyl)ether		1.454	1.662	1.509	1.473	1.413	1.502	6.4
Dimethylphthalate	1.341	1.334	1.504	1.396	1.245		1.364	7.0
2,6-Dinitrotoluene	0.217	0.238	0.277	0.269	0.265		0.253	9.9
Acenaphthylene	1.652	1.643	1.868	1.677	1.477		1.663	8.3
3-Nitroaniline		0.221	0.250	0.240	0.228	0.232	0.234	4.7
Acenaphthene	1.155	1.191	1.295	1.175	1.079		1.179	6.6
2,4-Dinitrophenol		0.074	0.104	0.128	0.162	0.179	0.129	32.8
4-Nitrophenol		0.151	0.194	0.213	0.216	0.228	0.200	15.0
Dibenzofuran	1.605	1.682	1.847	1.695	1.482		1.662	8.0
2,4-Dinitrotoluene	0.334	0.346	0.411	0.406	0.384		0.376	9.3
Diethylphthalate	1.290	1.289	1.462	1.338	1.214		1.319	7.0
2-Chlorophenol	0.995	1.193	1.372	1.249	1.227		1.207	11.3

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CHEM02
Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG No.: P5369
Instrument ID: BNA_G Calibration Date(s): 12/11/2024 12/11/2024
Calibration Time(s): 13:23 17:52

LAB FILE ID:		RRFAL1 = BG063651.D			RRFAL2 = BG063652.D		RRFAL3 = BG063653.D	
		RRFAL4 = BG063654.D			RRFAL5 = BG063655.D		RRFAL6 = BG063656.D	
COMPOUND	RRFAL1	RRFAL2	RRFAL3	RRFAL4	RRFAL5	RRFAL6	RRF	% RSD
Fluorene	1.338	1.354	1.472	1.310	1.163		1.327	8.3
4-Chlorophenyl-phenylether	0.737	0.741	0.814	0.743	0.663		0.739	7.2
4-Nitroaniline		0.208	0.267	0.248	0.224	0.226	0.234	9.8
4,6-Dinitro-2-methylphenol		0.087	0.106	0.112	0.115	0.115	0.107	11.1
N-Nitrosodiphenylamine	0.451	0.488	0.533	0.486	0.463		0.484	6.5
1,2,4,5-Tetrachlorobenzene	0.604	0.627	0.694	0.654	0.623		0.640	5.4
4-Bromophenyl-phenylether	0.187	0.198	0.227	0.207	0.200		0.204	7.4
Hexachlorobenzene	0.223	0.230	0.249	0.221	0.219		0.228	5.4
Atrazine		0.201	0.225	0.213	0.199	0.186	0.205	7.2
Pentachlorophenol		0.071	0.090	0.097	0.108	0.115	0.096	17.8
2-Methylphenol		1.316	1.568	1.465	1.384	1.361	1.419	7.0
Phenanthrene	0.964	0.985	1.066	0.970	0.834		0.964	8.6
Anthracene	0.958	0.994	1.104	0.987	0.835		0.975	9.9
Carbazole		0.822	0.925	0.848	0.754	0.599	0.790	15.6
Di-n-butylphthalate	0.957	0.964	1.072	0.956	0.820		0.954	9.4
Fluoranthene	1.179	1.254	1.324	1.190	0.967		1.183	11.3
Pyrene	1.265	1.339	1.449	1.252	0.987		1.259	13.6
Butylbenzylphthalate	0.382	0.370	0.415	0.390	0.375		0.386	4.5
3,3-Dichlorobenzidine		0.365	0.405	0.371	0.355	0.309	0.361	9.6
2,2-oxybis(1-Chloropropane)		2.203	2.500	2.268	2.137	2.040	2.230	7.8
Benzo(a)anthracene	1.254	1.274	1.415	1.251	1.073		1.253	9.7
Chrysene	1.228	1.201	1.351	1.189	1.004		1.195	10.4
Bis(2-ethylhexyl)phthalate	0.563	0.550	0.617	0.559	0.521		0.562	6.3
Di-n-octyl phthalate		0.914	0.987	0.912	0.826	0.677	0.863	13.8
Benzo(b)fluoranthene	1.064	1.121	1.270	1.150	1.079		1.137	7.2
Benzo(k)fluoranthene	1.090	1.149	1.279	1.148	1.003		1.134	8.9
Benzo(a)pyrene	0.989	1.046	1.199	1.094	1.007		1.067	7.9
Indeno(1,2,3-cd)pyrene	1.193	1.272	1.436	1.325	1.290		1.303	6.8
Dibenzo(a,h)anthracene	0.930	1.020	1.169	1.066	1.028		1.043	8.3
Benzo(g,h,i)perylene	0.977	1.012	1.133	1.041	1.007		1.034	5.8
Acetophenone		2.390	2.742	2.504	2.332	2.160	2.426	8.9
2,3,4,6-Tetrachlorophenol	0.275	0.320	0.346	0.348	0.339		0.326	9.2
1,4-Dioxane-d8	0.446	0.484	0.598	0.481	0.543		0.510	11.8
Pyridine-d5		1.381	1.735	1.588	1.794	1.553	1.610	10.1
Phenol-d5		1.679	2.009	1.888	1.812	1.823	1.842	6.5

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CHEM02
Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG No.: P5369
Instrument ID: BNA_G Calibration Date(s): 12/11/2024 12/11/2024
Calibration Time(s): 13:23 17:52

LAB FILE ID:		RRFAL1 = BG063651.D			RRFAL2 = BG063652.D		RRFAL3 = BG063653.D	
		RRFAL4 = BG063654.D			RRFAL5 = BG063655.D		RRFAL6 = BG063656.D	
COMPOUND	RRFAL1	RRFAL2	RRFAL3	RRFAL4	RRFAL5	RRFAL6	RRF	% RSD
Bis-(2-Chloroethyl) ether-d8		1.201	1.404	1.260	1.261	1.236	1.272	6.1
2-Chlorophenol-d4	1.063	1.162	1.331	1.232	1.175		1.193	8.2
4-Methylphenol-d8		1.365	1.553	1.475	1.372	1.386	1.430	5.7
Nitrobenzene-d5	0.125	0.139	0.153	0.149	0.148		0.143	7.7
2-Nitrophenol-d4	0.149	0.145	0.169	0.171	0.168		0.160	7.6
2,4-Dichlorophenol-d3	0.275	0.296	0.343	0.324	0.312		0.310	8.4
4-Chloroaniline-d4		0.377	0.422	0.414	0.391	0.364	0.394	6.2
4-Methylphenol		1.415	1.728	1.625	1.514	1.482	1.553	8.0
Dimethylphthalate-d6	1.331	1.366	1.489	1.367	1.251		1.361	6.3
Acenaphthylene-d8	1.521	1.520	1.726	1.572	1.418		1.551	7.2
4-Nitrophenol-d4		0.145	0.195	0.204	0.207	0.213	0.193	14.4
Fluorene-d10	1.172	1.190	1.335	1.215	1.104		1.203	7.0
4,6-Dinitro-2-methylphenol		0.088	0.099	0.107	0.111	0.113	0.104	9.9
Anthracene-d10	0.869	0.871	0.938	0.848	0.773		0.860	6.9
Pyrene-d10	1.011	1.102	1.159	1.044	0.887		1.041	9.9
Benzo(a)pyrene-d12	0.866	0.913	1.046	0.956	0.916		0.939	7.2
N-Nitroso-di-n-propylamine	1.342	1.371	1.614	1.522	1.388		1.447	8.0

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1