

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG NO.: P5369

EPA Sample No.: SSTD020584 Date Analyzed: 12/26/2024

Lab File ID: BG063845.D Time Analyzed: 22:02

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	88227	7.798	358736	10.59	277241	14.44
UPPER LIMIT	176454	8.298	717472	11.088	554482	14.937
LOWER LIMIT	44113.5	7.298	179368	10.088	138621	13.937
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	86844	7.80	337696	10.59	267806	14.44

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.