

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG NO.: P5369

EPA Sample No.: SSTD020584 Date Analyzed: 12/26/2024

Lab File ID: BG063845.D Time Analyzed: 22:02

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	683480	17.192	711419	21.44	747668	24.46
UPPER LIMIT	1366960	17.692	1422840	21.94	1495340	24.96
LOWER LIMIT	341740	16.692	355710	20.94	373834	23.96
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	697595	17.19	743341	21.45	785943	24.47

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.