

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG NO.: P5369
 Lab File ID: VN085339.D Date Analyzed: 12/30/2024
 Instrument ID: MSVOA_N Time Analyzed: 12:27
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	129301	13.788				
UPPER LIMIT	258602	14.288				
LOWER LIMIT	64650.5	13.288				
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	74220	13.79				
STORAGE-BLANK-SOIL-REF-5	86105	13.79				
STORAGE-BLANK-SOIL-REF-4	86818	13.79				
STORAGE-BLANK-SAMPLE-MANAGEMENT	83413	13.79				
VN1230WBL01	97891	13.79				
VN1230WBS01	121453	13.79				
VN1230WBSD01	117451	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.