

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CHEM02
Lab Code: CHEM Case No.: P5369 SAS No.: P5369 SDG NO.: P5369
Lab File ID: VN085320.D BFB Injection Date: 12/27/2024
Instrument ID: MSVOA_N BFB Injection Time: 09:52
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.4
75	30.0 - 60.0% of mass 95	58
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.3
173	Less than 2.0% of mass 174	0.6 (0.8) 1
174	50.0 - 100.0% of mass 95	73.9
175	5.0 - 9.0% of mass 174	5.8 (7.9) 1
176	95.0 - 101.0% of mass 174	73.2 (99) 1
177	5.0 - 9.0% of mass 176	4.8 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC100	VSTDICC100	VN085321.D	12/27/2024	10:26
VSTDICCC050	VSTDICCC050	VN085322.D	12/27/2024	10:50
VSTDICC020	VSTDICC020	VN085323.D	12/27/2024	11:13
VSTDICC010	VSTDICC010	VN085324.D	12/27/2024	11:37
VSTDICC005	VSTDICC005	VN085325.D	12/27/2024	12:01
VSTDICC001	VSTDICC001	VN085326.D	12/27/2024	12:48