



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Summit Environmental Technologies, LLC

Project Name: 24120686

Project # N/A

Chemtech Project # P5370

Test Name: TOC

A. Number of Samples and Date of Receipt:

2 Water samples were received on 12/20/2024.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: TOC. This data package contains results for TOC.

C. Analytical Techniques:

The analysis of TOC was based on method SM5310B.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike (120524-89MS) analysis met criteria for all samples except for TOC due to matrix interferences.

The Matrix Spike Duplicate (120524-89MSD) analysis met criteria for all samples except for TOC due to matrix interferences.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

P5370-01 and P5370-02 samples were analyzed straight for TOC Due to bad matrix, analyzed as 400X straight dilution.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____