

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: P5371 SAS No.: P5371 SDG No.: P5371

Instrument ID: MSVOA_N Calibration Date/Time: 12/23/2024 11:25

Lab File ID: VN085286.D Init. Calib. Date(s): 12/18/2024 12/18/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:56 13:57

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.797	0.836		4.89	20
Chloromethane	0.759	0.805	0.1	6.06	20
Vinyl Chloride	0.711	0.810		13.92	20
Bromomethane	0.467	0.469		0.43	20
Chloroethane	0.513	0.476		-7.21	20
Trichlorofluoromethane	1.149	1.150		0.09	20
1,1,2-Trichlorotrifluoroethane	0.616	0.665		7.95	20
1,1-Dichloroethene	0.571	0.610		6.83	20
Acetone	0.302	0.247		-18.21	20
Carbon Disulfide	1.948	1.877		-3.64	20
Methyl tert-butyl Ether	1.956	1.915		-2.1	20
Methyl Acetate	0.854	0.695		-18.62	20
Methylene Chloride	0.660	0.672		1.82	20
trans-1,2-Dichloroethene	0.629	0.631		0.32	20
1,1-Dichloroethane	1.221	1.284	0.1	5.16	20
Cyclohexane	1.131	1.102		-2.56	20
2-Butanone	0.462	0.359		-22.29	20
Carbon Tetrachloride	0.592	0.592		0	20
cis-1,2-Dichloroethene	0.733	0.742		1.23	20
Bromochloromethane	0.616	0.633		2.76	20
Chloroform	1.304	1.269		-2.68	20
1,1,1-Trichloroethane	1.162	1.127		-3.01	20
Methylcyclohexane	0.509	0.611		20.04	20
Benzene	1.523	1.661		9.06	20
1,2-Dichloroethane	0.580	0.567		-2.24	20
Trichloroethene	0.339	0.379		11.8	20
1,2-Dichloropropane	0.377	0.414		9.81	20
Bromodichloromethane	0.577	0.589		2.08	20
4-Methyl-2-Pentanone	0.543	0.443		-18.42	20
Toluene	0.916	0.994		8.52	20
t-1,3-Dichloropropene	0.562	0.595		5.87	20
cis-1,3-Dichloropropene	0.598	0.653		9.2	20
1,1,2-Trichloroethane	0.343	0.351		2.33	20
2-Hexanone	0.387	0.312		-19.38	20
Dibromochloromethane	0.406	0.412		1.48	20
1,2-Dibromoethane	0.353	0.342		-3.12	20
Tetrachloroethene	0.336	0.384		14.29	20
Chlorobenzene	1.124	1.200	0.3	6.76	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: P5371 SAS No.: P5371 SDG No.: P5371

Instrument ID: MSVOA_N Calibration Date/Time: 12/23/2024 11:25

Lab File ID: VN085286.D Init. Calib. Date(s): 12/18/2024 12/18/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:56 13:57

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.926	2.142		11.22	20
m/p-Xylenes	0.721	0.819		13.59	20
o-Xylene	0.677	0.764		12.85	20
Styrene	1.131	1.293		14.32	20
Bromoform	0.307	0.284	0.1	-7.49	20
Isopropylbenzene	3.607	4.148		15	20
1,1,2,2-Tetrachloroethane	1.285	1.113	0.3	-13.39	20
1,3-Dichlorobenzene	1.744	1.836		5.28	20
1,4-Dichlorobenzene	1.805	1.821		0.89	20
1,2-Dichlorobenzene	1.622	1.742		7.4	20
1,2-Dibromo-3-Chloropropane	0.246	0.199		-19.11	20
1,2,4-Trichlorobenzene	0.830	0.915		10.24	20
1,2,3-Trichlorobenzene	0.829	0.849		2.41	20
1,2-Dichloroethane-d4	0.874	0.769		-12.01	20
Dibromofluoromethane	0.352	0.344		-2.27	20
Toluene-d8	1.285	1.305		1.56	20
4-Bromofluorobenzene	0.459	0.428		-6.75	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.