

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: P5371 SAS No.: P5371 SDG NO.: P5371
 Lab File ID: VN085286.D Date Analyzed: 12/23/2024
 Instrument ID: MSVOA_N Time Analyzed: 11:25
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	177949	8.22	299262	9.09	263260	11.86
UPPER LIMIT	355898	8.718	598524	9.594	526520	12.359
LOWER LIMIT	88974.5	7.718	149631	8.594	131630	11.359
EPA SAMPLE NO.						
SPLP-C10-1-M	171305	8.22	315145	9.09	279146	11.86
SPLP-C10-1-MDL	129978	8.22	242088	9.10	214104	11.87
SPLP-C10-2-M	197251	8.22	350073	9.09	303609	11.86
SPLP-C10-2-MDL	152849	8.22	276135	9.09	242482	11.86
SPLP-C10-3-M	204753	8.22	354047	9.09	317295	11.86
SPLP-C10-3-MDL	153204	8.22	286168	9.09	239567	11.86
SPLP-C11-1-M	190171	8.22	333838	9.10	292239	11.86
SPLP-C11-1-MDL	147336	8.22	268390	9.10	235244	11.86
SPLP-C11-2-M	181620	8.22	322846	9.09	281172	11.86
SPLP-C11-3-M	177938	8.22	313792	9.10	274059	11.86
SPLP-C12-1-M	167201	8.22	305341	9.09	263198	11.87
SPLP-C12-2-M	162897	8.22	294714	9.09	257534	11.87
SPLP-C12-3-M	160117	8.22	295060	9.09	257975	11.86
VN1223WBL01	174137	8.22	324722	9.09	277455	11.86
VN1223WBS02	188847	8.22	320551	9.09	275426	11.86
VN1223WBSD02	146303	8.22	258425	9.09	217793	11.86

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: P5371 SAS No.: P5371 SDG NO.: P5371
 Lab File ID: VN085286.D Date Analyzed: 12/23/2024
 Instrument ID: MSVOA_N Time Analyzed: 11:25
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	177949	8.22	299262	9.09	263260	11.86
UPPER LIMIT	355898	8.718	598524	9.594	526520	12.359
LOWER LIMIT	88974.5	7.718	149631	8.594	131630	11.359
EPA SAMPLE NO.						

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.