

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05  
 Lab Code: CHEM Case No.: P5371 SAS No.: P5371 SDG NO.: P5371  
 Lab File ID: VN085309.D Date Analyzed: 12/26/2024  
 Instrument ID: MSVOA\_N Time Analyzed: 10:10  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 159033        | 8.22  | 279635        | 9.09  | 241438        | 11.86  |
| UPPER LIMIT    | 318066        | 8.718 | 559270        | 9.594 | 482876        | 12.359 |
| LOWER LIMIT    | 79516.5       | 7.718 | 139818        | 8.594 | 120719        | 11.359 |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| SPLP-C11-2-MDL | 132134        | 8.22  | 242328        | 9.10  | 206929        | 11.86  |
| SPLP-C11-3-MRE | 152126        | 8.22  | 275786        | 9.10  | 242276        | 11.87  |
| SPLP-C12-1-MRE | 153657        | 8.22  | 277699        | 9.10  | 236786        | 11.86  |
| SPLP-C12-2-MRE | 149548        | 8.22  | 279987        | 9.10  | 247264        | 11.87  |
| SPLP-C12-3-MRE | 149058        | 8.22  | 273099        | 9.10  | 239856        | 11.87  |
| VN1226WBL01    | 159396        | 8.22  | 294503        | 9.09  | 251735        | 11.86  |
| VN1226WBS01    | 145338        | 8.22  | 258792        | 9.10  | 222802        | 11.86  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.