

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: P5371 SAS No.: P5371 SDG NO.: P5371
 Lab File ID: VN085309.D Date Analyzed: 12/26/2024
 Instrument ID: MSVOA_N Time Analyzed: 10:10
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	111930	13.788				
UPPER LIMIT	223860	14.288				
LOWER LIMIT	55965	13.288				
EPA SAMPLE NO.						
SPLP-C11-2-MDL	87889	13.79				
SPLP-C11-3-MRE	109298	13.79				
SPLP-C12-1-MRE	104487	13.79				
SPLP-C12-2-MRE	114087	13.79				
SPLP-C12-3-MRE	110672	13.79				
VN1226WBL01	102404	13.79				
VN1226WBS01	98856	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.