ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092 NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEM	ITECH PROJECT NUMBER: P5374 MATRI	IX: Water		
METH	IOD: 8270-Modified/3510			
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Met	NA	NO	YES ✓
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CI CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 8000 Series.	12 hours for		✓
4.	GC/MS Calibration - Initial Calibration performed within 30 days before analysis and continuing calibration performed within 24 hours of sample 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements . The Continuous Calibration File ID BN035811.D met the requirements et 2,4,6-Tribromophenol, The failure compound not associated with the clie		✓	
	list, therefore no corrective action was taken. The Continuous Calibration File ID BN035845.D met the requirements e Nitrobenzene-d5, The failure compound not associated with the client par therefore no corrective action was taken.			
6.	Blank Contamination - If yes, list compounds and concentrations in each	blank:	\checkmark	
7.	Surrogate Recoveries Meet Criteria		\checkmark	
	If not met, list those compounds and their recoveries which fall outside the ranges.	ie acceptable		
	The Surrogate recoveries met the acceptable criteria except for PB165828 [Nitrobenzene-d5 - 122%], The failure surrogates not associated with the parameters list, therefore no corrective action was taken.			

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

<u>GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY</u> (CONTINUED)

		NA	NO	YES
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria		\checkmark	
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			
	The MS {P5376-08MS} with File ID: BN035821.D recoveries met the requirements for all compounds except for 1,4-Dioxane[53%], Due to a poor compound the recovery no corrective action required. The MSD {P5376-09MSD} with File ID: BN035822.D recoveries met the acceptable requirements except for 1,4-Dioxane[51%], Due to a poor compound the recovery no corrective action required. The Blank Spike met requirements for all samples .			
9.	Internal Standard Area/Retention Time Shift Meet Criteria Comments:			\checkmark
10.	Extraction Holding Time Met			\checkmark
	If not met, list number of days exceeded for each sample:			
11.	Analysis Holding Time Met			\checkmark
	If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)." The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points. The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW