

Surrogate Summary

SW-846

SDG No.: P5374

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P5374-03	BP-VPB-190A-GW-718-720	2-Methylnaphthalene-d10	0.4	0.28	70		30	150
		Fluoranthene-d10	0.4	0.22	55		30	150
		Nitrobenzene-d5	0.4	0.24	60		55	111
		2-Fluorobiphenyl	0.4	0.21	53		53	106
		Terphenyl-d14	0.4	0.29	73		58	132
P5376-08MS	RW10-MW01S-20241218MS	2-Methylnaphthalene-d10	0.4	0.43	107		30	150
		Fluoranthene-d10	0.4	0.43	106		30	150
		Nitrobenzene-d5	0.4	0.42	106		55	111
		2-Fluorobiphenyl	0.4	0.38	94		53	106
		Terphenyl-d14	0.4	0.49	122		58	132
P5376-09MSD	RW10-MW01S-20241218MSD	2-Methylnaphthalene-d10	0.4	0.43	108		30	150
		Fluoranthene-d10	0.4	0.43	108		30	150
		Nitrobenzene-d5	0.4	0.42	104		55	111
		2-Fluorobiphenyl	0.4	0.37	93		53	106
		Terphenyl-d14	0.4	0.51	128		58	132
PB165828BL	PB165828BL	2-Methylnaphthalene-d10	0.4	0.34	86		30	150
		Fluoranthene-d10	0.4	0.40	100		30	150
		Nitrobenzene-d5	0.4	0.42	105		55	111
		2-Fluorobiphenyl	0.4	0.34	86		53	106
		Terphenyl-d14	0.4	0.48	120		58	132
PB165828BS	PB165828BS	2-Methylnaphthalene-d10	0.4	0.47	116		30	150
		Fluoranthene-d10	0.4	0.34	85		30	150
		Nitrobenzene-d5	0.4	0.49	122	*	55	111
		2-Fluorobiphenyl	0.4	0.39	98		53	106
		Terphenyl-d14	0.4	0.48	119		58	132