

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5374 SAS No.: P5374 SDG NO.: P5374

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 12/26/2024

Lab File ID: BN035845.D Time Analyzed: 09:58

Instrument ID: BNA\_N GC Column: ZB-GR ID: 0.25 (mm)

|                           | IS1 (DCB)<br>AREA # | RT #  | IS2 (NPT)<br>AREA # | RT #   | IS3 (ANT)<br>AREA # | RT #   |
|---------------------------|---------------------|-------|---------------------|--------|---------------------|--------|
| 12 HOUR STD               | 3755                | 7.257 | 9322                | 10.00  | 5387                | 13.91  |
| UPPER LIMIT               | 7510                | 7.757 | 18644               | 10.499 | 10774               | 14.414 |
| LOWER LIMIT               | 1877.5              | 6.757 | 4661                | 9.499  | 2693.5              | 13.414 |
| EPA SAMPLE NO.            |                     |       |                     |        |                     |        |
| 01 BP-VPB-190A-GW-718-720 | 3717                | 7.26  | 9841                | 10.00  | 7032                | 13.91  |
| 02 PB165828BS             | 4431                | 7.26  | 10905               | 10.00  | 6427                | 13.91  |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.